

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE070617\
 Data File : BE093427.D
 Acq On : 6 Jul 2017 10:29
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 07 06:48:49 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE070217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 03 06:40:23 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
2	1,4-Dioxane	0.847	0.782	7.7	94	-0.01
3	n-Nitrosodimethylamine	0.444	0.446	-0.5	128	0.00
4 S	2-Fluorophenol	1.207	1.216	-0.7	105	0.00
5 S	Phenol-d6	1.780	1.911	-7.4	116	0.00
6	bis(2-Chloroethyl)ether	1.339	1.292	3.5	100	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	112	0.00
8 S	Nitrobenzene-d5	0.256	0.283	-10.5	134	0.00
9	Naphthalene	4.190	4.064	3.0	112	0.00
10	Hexachlorobutadiene	0.159	0.157	1.3	112	0.00
11 SURR	2-Methylnaphthalene-d10	0.638	0.634	0.6	115	0.00
12	2-Methylnaphthalene	0.704	0.695	1.3	115	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	111	0.00
14 S	2,4,6-Tribromophenol	0.126	0.128	-1.6	124	0.00
15 S	2-Fluorobiphenyl	1.553	1.510	2.8	112	0.00
16	Acenaphthylene	1.666	1.661	0.3	117	0.00
17	Acenaphthene	1.220	1.179	3.4	112	0.00
18	Fluorene	1.675	1.558	7.0	107	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	104	0.00
20	4-Bromophenyl-phenylether	0.219	0.187	14.6	96	0.00
21	Hexachlorobenzene	0.237	0.226	4.6	110	0.00
22	Pentachlorophenol	0.035	0.028	20.0	102	0.00
23	Phenanthrene	1.454	1.345	7.5	105	0.00
24	Anthracene	0.891	0.794	10.9	98	0.00
25 SURR	Fluoranthene-d10	4.823	4.257	11.7	100	0.00
26	Fluoranthene	1.367	1.212	11.3	101	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	99	0.00
28	Pyrene	1.124	1.101	2.0	99	0.00
29 S	Terphenyl-d14	0.714	0.693	2.9	98	0.00
30	Benzo(a)anthracene	1.080	1.109	-2.7	107	0.00
31	Chrysene	1.130	1.099	2.7	98	0.00
32	Bis(2-ethylhexyl)phthalate	1.555	1.841	-18.4	143	-0.01
33	Indeno(1,2,3-cd)pyrene	0.992	1.076	-8.5	111	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	110	0.00
35	Benzo(b)fluoranthene	1.292	1.180	8.7	103	0.00
36	Benzo(k)fluoranthene	1.278	1.195	6.5	108	0.00
37 C	Benzo(a)pyrene	1.132	1.086	4.1	111	0.00
38	Dibenzo(a,h)anthracene	1.099	1.063	3.3	110	0.00
39	Benzo(g,h,i)perylene	1.110	1.064	4.1	111	0.00

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE070617\
Data File : BE093427.D
Acq On : 6 Jul 2017 10:29
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 07 06:48:49 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE070217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 03 06:40:23 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			