

Data Path : Z:\HPCHEM1\BNA_E\Data\BE070617\
 Data File : BE093432.D
 Acq On : 6 Jul 2017 16:24
 Operator : SJ/JU
 Sample : SSTDC00.4
 Misc : LOQ 0.2 PPM
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDC00.4

Quant Time: Jul 07 07:08:39 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE070217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 03 06:40:23 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	127	0.00
2	1,4-Dioxane	0.847	0.791	6.6	118	0.00
3	n-Nitrosodimethylamine	0.444	0.440	0.9	158#	0.00
4 S	2-Fluorophenol	1.207	1.142	5.4	123	0.00
5 S	Phenol-d6	1.780	1.824	-2.5	139	0.00
6	bis(2-Chloroethyl)ether	1.339	1.303	2.7	126	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	142	0.00
8 S	Nitrobenzene-d5	0.256	0.277	-8.2	166#	0.00
9	Naphthalene	4.190	4.045	3.5	140	0.00
10	Hexachlorobutadiene	0.159	0.150	5.7	136	0.00
11 SURR	2-Methylnaphthalene-d10	0.638	0.639	-0.2	147	0.00
12	2-Methylnaphthalene	0.704	0.700	0.6	146	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	141	0.00
14 S	2,4,6-Tribromophenol	0.126	0.107	15.1	132	0.00
15 S	2-Fluorobiphenyl	1.553	1.551	0.1	146	0.00
16	Acenaphthylene	1.666	1.604	3.7	144	0.00
17	Acenaphthene	1.220	1.176	3.6	142	0.00
18	Fluorene	1.675	1.541	8.0	135	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	121	0.00
20	4-Bromophenyl-phenylether	0.219	0.222	-1.4	132	0.00
21	Hexachlorobenzene	0.237	0.233	1.7	133	0.00
22	Pentachlorophenol	0.035	0.028	20.0	117	0.00
23	Phenanthrene	1.454	1.350	7.2	122	0.00
24	Anthracene	0.891	0.714	19.9	103	0.00
25 SURR	Fluoranthene-d10	4.823	3.980	17.5	109	0.00
26	Fluoranthene	1.367	1.139	16.7	110	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	102	0.00
28	Pyrene	1.124	1.157	-2.9	108	0.00
29 S	Terphenyl-d14	0.714	0.724	-1.4	105	0.00
30	Benzo(a)anthracene	1.080	1.043	3.4	104	0.00
31	Chrysene	1.130	1.098	2.8	101	0.00
32	Bis(2-ethylhexyl)phthalate	1.555	1.502	3.4	120	-0.01
33	Indeno(1,2,3-cd)pyrene	0.992	1.042	-5.0	110	0.00
34 I	Perylene-d12	1.000	1.000	0.0	108	0.00
35	Benzo(b)fluoranthene	1.292	1.199	7.2	103	0.00
36	Benzo(k)fluoranthene	1.278	1.248	2.3	111	0.00
37 C	Benzo(a)pyrene	1.132	1.096	3.2	111	0.00
38	Dibenzo(a,h)anthracene	1.099	1.073	2.4	110	0.00
39	Benzo(g,h,i)perylene	1.110	1.058	4.7	108	0.00

Data Path : Z:\HPCHEM1\BNA_E\Data\BE070617\
Data File : BE093432.D
Acq On : 6 Jul 2017 16:24
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc : LOQ 0.2 PPM
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 07 07:08:39 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE070217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 03 06:40:23 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			