

Data Path : Z:\HPCHEM1\BNA_E\Data\BE070717\
 Data File : BE093434.D
 Acq On : 7 Jul 2017 8:55
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 08 04:50:18 2017
 Quant Title :
 QLast Update : Mon Jul 03 06:40:23 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	114	0.00
2	1,4-Dioxane	0.400	0.392	2.0	112	0.00
3	n-Nitrosodimethylamine	0.400	0.427	-6.7	151	0.00
4 S	2-Fluorophenol	0.400	0.415	-3.7	121	0.00
5 S	Phenol-d6	0.400	0.425	-6.2	129	0.00
6	bis(2-Chloroethyl)ether	0.400	0.426	-6.5	124	-0.01
7 I	Naphthalene-d8	0.400	0.400	0.0	122	0.00
8 S	Nitrobenzene-d5	0.400	0.430	-7.5	142	0.00
9	Naphthalene	0.400	0.417	-4.2	130	0.00
10	Hexachlorobutadiene	0.400	0.416	-4.0	129	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.419	-4.7	132	0.00
12	2-Methylnaphthalene	0.400	0.417	-4.2	132	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	119	0.00
14 S	2,4,6-Tribromophenol	0.400	0.357	10.8	118	0.00
15 S	2-Fluorobiphenyl	0.400	0.418	-4.5	130	0.00
16	Acenaphthylene	0.400	0.402	-0.5	127	0.00
17	Acenaphthene	0.400	0.413	-3.2	129	0.00
18	Fluorene	0.400	0.394	1.5	122	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	97	0.00
20	4-Bromophenyl-phenylether	0.400	0.451	-12.7	118	0.00
21	Hexachlorobenzene	0.400	0.456	-14.0	123	0.00
22	Pentachlorophenol	0.400	0.430	-7.5	111	0.00
23	Phenanthrene	0.400	0.430	-7.5	114	0.00
24	Anthracene	0.400	0.406	-1.5	104	0.00
25 SURR	Fluoranthene-d10	0.400	0.405	-1.3	107	0.00
26	Fluoranthene	0.400	0.404	-1.0	108	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	98	0.00
28	Pyrene	0.400	0.426	-6.5	106	0.00
29 S	Terphenyl-d14	0.400	0.429	-7.2	107	0.00
30	Benzo(a)anthracene	0.400	0.425	-6.2	109	0.00
31	Chrysene	0.400	0.415	-3.7	103	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.414	-3.5	123	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.444	-11.0	112	-0.01
34 I	Perylene-d12	0.400	0.400	0.0	103	0.00
35	Benzo(b)fluoranthene	0.400	0.395	1.3	104	0.00
36	Benzo(k)fluoranthene	0.400	0.416	-4.0	113	0.00
37 C	Benzo(a)pyrene	0.400	0.409	-2.2	111	0.00
38	Dibenzo(a,h)anthracene	0.400	0.413	-3.2	111	-0.01
39	Benzo(g,h,i)perylene	0.400	0.412	-3.0	112	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0