

Data Path : Z:\HPCHEM1\BNA_E\Data\BE070717\
 Data File : BE093434.D
 Acq On : 7 Jul 2017 8:55
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 08 04:50:18 2017
 Quant Title :
 QLast Update : Mon Jul 03 06:40:23 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	114	0.00
2	1,4-Dioxane	0.847	0.831	1.9	112	0.00
3	n-Nitrosodimethylamine	0.444	0.468	-5.4	151#	0.00
4 S	2-Fluorophenol	1.207	1.251	-3.6	121	0.00
5 S	Phenol-d6	1.780	1.893	-6.3	129	0.00
6	bis(2-Chloroethyl)ether	1.339	1.427	-6.6	124	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	122	0.00
8 S	Nitrobenzene-d5	0.256	0.275	-7.4	142	0.00
9	Naphthalene	4.190	4.363	-4.1	130	0.00
10	Hexachlorobutadiene	0.159	0.165	-3.8	129	0.00
11 SURR	2-Methylnaphthalene-d10	0.638	0.668	-4.7	132	0.00
12	2-Methylnaphthalene	0.704	0.734	-4.3	132	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	119	0.00
14 S	2,4,6-Tribromophenol	0.126	0.112	11.1	118	0.00
15 S	2-Fluorobiphenyl	1.553	1.624	-4.6	130	0.00
16	Acenaphthylene	1.666	1.676	-0.6	127	0.00
17	Acenaphthene	1.220	1.259	-3.2	129	0.00
18	Fluorene	1.675	1.651	1.4	122	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	97	0.00
20	4-Bromophenyl-phenylether	0.219	0.247	-12.8	118	0.00
21	Hexachlorobenzene	0.237	0.270	-13.9	123	0.00
22	Pentachlorophenol	0.035	0.033	5.7	111	0.00
23	Phenanthrene	1.454	1.563	-7.5	114	0.00
24	Anthracene	0.891	0.905	-1.6	104	0.00
25 SURR	Fluoranthene-d10	4.823	4.881	-1.2	107	0.00
26	Fluoranthene	1.367	1.382	-1.1	108	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	98	0.00
28	Pyrene	1.124	1.196	-6.4	106	0.00
29 S	Terphenyl-d14	0.714	0.767	-7.4	107	0.00
30	Benzo(a)anthracene	1.080	1.147	-6.2	109	0.00
31	Chrysene	1.130	1.171	-3.6	103	0.00
32	Bis(2-ethylhexyl)phthalate	1.555	1.608	-3.4	123	-0.01
33	Indeno(1,2,3-cd)pyrene	0.992	1.102	-11.1	112	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	103	0.00
35	Benzo(b)fluoranthene	1.292	1.275	1.3	104	0.00
36	Benzo(k)fluoranthene	1.278	1.331	-4.1	113	0.00
37 C	Benzo(a)pyrene	1.132	1.156	-2.1	111	0.00
38	Dibenzo(a,h)anthracene	1.099	1.136	-3.4	111	-0.01
39	Benzo(g,h,i)perylene	1.110	1.143	-3.0	112	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0