

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE070815\
 Data File : BE090062.D
 Acq On : 8 Jul 2015 16:51
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Jul 09 01:19:06 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BE062915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 30 07:01:01 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	138	-0.01
2	1,4-Dioxane	0.686	0.580	15.5	121	-0.01
3	n-Nitrosodimethylamine	0.930	0.766	17.6	118	-0.02
4 S	2-Fluorophenol	1.149	0.984	14.4	118	-0.01
5 S	Phenol-d6	1.608	1.448	10.0	126	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	140	-0.01
7 S	Nitrobenzene-d5	0.397	0.344	13.4	123	-0.01
8	Nitrobenzene	0.421	0.361	14.3	124	-0.01
9	Naphthalene	1.130	1.038	8.1	131	-0.01
10	Hexachlorobutadiene	0.178	0.166	6.7	133	-0.01
11	2-Methylnaphthalene	0.755	0.667	11.7	126	-0.01
12 I	Acenaphthene-d10	1.000	1.000	0.0	132	-0.01
13 S	2,4,6-Tribromophenol	0.147	0.098	33.3#	90	0.00
14 S	2-Fluorobiphenyl	1.503	1.356	9.8	122	-0.01
15	Acenaphthylene	9.019	8.220	8.9	124	-0.01
16	Acenaphthene	1.300	1.207	7.2	124	-0.01
17	Fluorene	1.768	1.578	10.7	121	-0.02
18 I	Phenanthrene-d10	1.000	1.000	0.0	122	-0.02
19	4-Bromophenyl-phenylether	0.213	0.187	12.2	110	-0.02
20	Hexachlorobenzene	0.885	0.785	11.3	113	-0.02
21	Pentachlorophenol	0.086	0.044	48.8#	67	0.00
22	Phenanthrene	1.281	1.120	12.6	110	-0.02
23	Anthracene	1.172	1.045	10.8	111	-0.02
24	Fluoranthene	1.431	1.249	12.7	109	-0.02
25 I	Chrysene-d12	1.000	1.000	0.0	113	-0.01
26	Pyrene	1.192	1.143	4.1	111	-0.02
27 S	Terphenyl-d14	0.833	0.762	8.5	104	-0.02
28	Benzo(a)anthracene	1.249	1.116	10.6	104	-0.01
29	Chrysene	1.297	1.169	9.9	104	-0.01
30	Bis(2-ethylhexyl)phthalate	0.398	0.242	39.2#	75	-0.01
31	Indeno(1,2,3-cd)pyrene	1.243	1.092	12.1	101	-0.04
32 I	Perylene-d12	1.000	1.000	0.0	104	-0.02
33	Benzo(b)fluoranthene	1.103	1.098	0.5	105	-0.02
34	Benzo(k)fluoranthene	1.232	1.245	-1.1	106	-0.02
35 C	Benzo(a)pyrene	1.019	1.000	1.9	103	-0.02
36	Dibenzo(a,h)anthracene	1.046	0.972	7.1	99	-0.03
37	Benzo(g,h,i)perylene	1.060	0.982	7.4	98	-0.04

(#) = Out of Range

SPCC's out = 0 CCC's out = 0