

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE070815\
 Data File : BE090062.D
 Acq On : 8 Jul 2015 16:51
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Jul 09 01:19:06 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BE062915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 30 07:01:01 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	138	-0.01
2	1,4-Dioxane	1.000	0.845	15.5	121	-0.01
3	n-Nitrosodimethylamine	1.000	0.824	17.6	118	-0.02
4 S	2-Fluorophenol	1.000	0.857	14.3	118	-0.01
5 S	Phenol-d6	1.000	0.901	9.9	126	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	140	-0.01
7 S	Nitrobenzene-d5	1.000	0.865	13.5	123	-0.01
8	Nitrobenzene	1.000	0.858	14.2	124	-0.01
9	Naphthalene	1.000	0.918	8.2	131	-0.01
10	Hexachlorobutadiene	1.000	0.935	6.5	133	-0.01
11	2-Methylnaphthalene	1.000	0.883	11.7	126	-0.01
12 I	Acenaphthene-d10	5.000	5.000	0.0	132	-0.01
13 S	2,4,6-Tribromophenol	1.000	0.667	33.3#	90	0.00
14 S	2-Fluorobiphenyl	1.000	0.902	9.8	122	-0.01
15	Acenaphthylene	1.000	0.911	8.9	124	-0.01
16	Acenaphthene	1.000	0.928	7.2	124	-0.01
17	Fluorene	1.000	0.892	10.8	121	-0.02
18 I	Phenanthrene-d10	5.000	5.000	0.0	122	-0.02
19	4-Bromophenyl-phenylether	1.000	0.880	12.0	110	-0.02
20	Hexachlorobenzene	1.000	0.888	11.2	113	-0.02
21	Pentachlorophenol	1.000	0.638	36.2#	67	0.00
22	Phenanthrene	1.000	0.874	12.6	110	-0.02
23	Anthracene	1.000	0.891	10.9	111	-0.02
24	Fluoranthene	1.000	0.873	12.7	109	-0.02
25 I	Chrysene-d12	5.000	5.000	0.0	113	-0.01
26	Pyrene	1.000	0.959	4.1	111	-0.02
27 S	Terphenyl-d14	1.000	0.914	8.6	104	-0.02
28	Benzo(a)anthracene	1.000	0.894	10.6	104	-0.01
29	Chrysene	1.000	0.902	9.8	104	-0.01
30	Bis(2-ethylhexyl)phthalate	1.000	0.702	29.8#	75	-0.01
31	Indeno(1,2,3-cd)pyrene	1.000	0.879	12.1	101	-0.04
32 I	Perylene-d12	5.000	5.000	0.0	104	-0.02
33	Benzo(b)fluoranthene	1.000	0.995	0.5	105	-0.02
34	Benzo(k)fluoranthene	1.000	1.010	-1.0	106	-0.02
35 C	Benzo(a)pyrene	1.000	0.982	1.8	103	-0.02
36	Dibenzo(a,h)anthracene	1.000	0.929	7.1	99	-0.03
37	Benzo(g,h,i)perylene	1.000	0.927	7.3	98	-0.04

(#) = Out of Range

SPCC's out = 0 CCC's out = 0