

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE070815\
 Data File : BE090074.D
 Acq On : 9 Jul 2015 00:34
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Jul 09 01:21:53 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BE062915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 30 07:01:01 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	147	0.00
2	1,4-Dioxane	0.686	0.522	23.9	116	0.00
3	n-Nitrosodimethylamine	0.930	0.755	18.8	124	0.00
4 S	2-Fluorophenol	1.149	1.167	-1.6	149	0.00
5 S	Phenol-d6	1.608	1.731	-7.6	161#	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	154#	-0.01
7 S	Nitrobenzene-d5	0.397	0.362	8.8	143	0.00
8	Nitrobenzene	0.421	0.373	11.4	140	0.00
9	Naphthalene	1.130	1.042	7.8	144	-0.01
10	Hexachlorobutadiene	0.178	0.167	6.2	146	0.00
11	2-Methylnaphthalene	0.755	0.672	11.0	139	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	139	0.00
13 S	2,4,6-Tribromophenol	0.147	0.159	-8.2	153#	0.00
14 S	2-Fluorobiphenyl	1.503	1.387	7.7	131	0.00
15	Acenaphthylene	9.019	8.569	5.0	136	0.00
16	Acenaphthene	1.300	1.226	5.7	133	-0.01
17	Fluorene	1.768	1.577	10.8	127	-0.02
18 I	Phenanthrene-d10	1.000	1.000	0.0	123	-0.02
19	4-Bromophenyl-phenylether	0.213	0.193	9.4	115	-0.02
20	Hexachlorobenzene	0.885	0.791	10.6	115	-0.02
21	Pentachlorophenol	0.086	0.085	1.2	129	0.00
22	Phenanthrene	1.281	1.164	9.1	116	-0.02
23	Anthracene	1.172	1.087	7.3	117	0.00
24	Fluoranthene	1.431	1.310	8.5	115	-0.01
25 I	Chrysene-d12	1.000	1.000	0.0	114	-0.01
26	Pyrene	1.192	1.193	-0.1	116	0.00
27 S	Terphenyl-d14	0.833	0.800	4.0	110	0.00
28	Benzo(a)anthracene	1.249	1.198	4.1	112	-0.01
29	Chrysene	1.297	1.191	8.2	106	-0.01
30	Bis(2-ethylhexyl)phthalate	0.398	0.602	-51.3#	188#	-0.01
31	Indeno(1,2,3-cd)pyrene	1.243	1.250	-0.6	117	-0.04
32 I	Perylene-d12	1.000	1.000	0.0	110	-0.02
33	Benzo(b)fluoranthene	1.103	1.185	-7.4	119	-0.02
34	Benzo(k)fluoranthene	1.232	1.257	-2.0	113	-0.02
35 C	Benzo(a)pyrene	1.019	1.101	-8.0	119	-0.02
36	Dibenzo(a,h)anthracene	1.046	1.063	-1.6	114	-0.03
37	Benzo(g,h,i)perylene	1.060	1.091	-2.9	115	-0.04

(#) = Out of Range

SPCC's out = 0 CCC's out = 0