

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE070815\
 Data File : BE090074.D
 Acq On : 9 Jul 2015 00:34
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Jul 09 01:21:53 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BE062915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 30 07:01:01 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	147	0.00
2	1,4-Dioxane	1.000	0.761	23.9	116	0.00
3	n-Nitrosodimethylamine	1.000	0.811	18.9	124	0.00
4 S	2-Fluorophenol	1.000	1.016	-1.6	149	0.00
5 S	Phenol-d6	1.000	1.077	-7.7	161	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	154	-0.01
7 S	Nitrobenzene-d5	1.000	0.911	8.9	143	0.00
8	Nitrobenzene	1.000	0.885	11.5	140	0.00
9	Naphthalene	1.000	0.921	7.9	144	-0.01
10	Hexachlorobutadiene	1.000	0.938	6.2	146	0.00
11	2-Methylnaphthalene	1.000	0.890	11.0	139	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	139	0.00
13 S	2,4,6-Tribromophenol	1.000	1.083	-8.3	153	0.00
14 S	2-Fluorobiphenyl	1.000	0.923	7.7	131	0.00
15	Acenaphthylene	1.000	0.950	5.0	136	0.00
16	Acenaphthene	1.000	0.944	5.6	133	-0.01
17	Fluorene	1.000	0.892	10.8	127	-0.02
18 I	Phenanthrene-d10	5.000	5.000	0.0	123	-0.02
19	4-Bromophenyl-phenylether	1.000	0.907	9.3	115	-0.02
20	Hexachlorobenzene	1.000	0.894	10.6	115	-0.02
21	Pentachlorophenol	1.000	0.984	1.6	129	0.00
22	Phenanthrene	1.000	0.909	9.1	116	-0.02
23	Anthracene	1.000	0.928	7.2	117	0.00
24	Fluoranthene	1.000	0.915	8.5	115	-0.01
25 I	Chrysene-d12	5.000	5.000	0.0	114	-0.01
26	Pyrene	1.000	1.001	-0.1	116	0.00
27 S	Terphenyl-d14	1.000	0.960	4.0	110	0.00
28	Benzo(a)anthracene	1.000	0.959	4.1	112	-0.01
29	Chrysene	1.000	0.918	8.2	106	-0.01
30	Bis(2-ethylhexyl)phthalate	1.000	1.290	-29.0#	188	-0.01
31	Indeno(1,2,3-cd)pyrene	1.000	1.006	-0.6	117	-0.04
32 I	Perylene-d12	5.000	5.000	0.0	110	-0.02
33	Benzo(b)fluoranthene	1.000	1.074	-7.4	119	-0.02
34	Benzo(k)fluoranthene	1.000	1.020	-2.0	113	-0.02
35 C	Benzo(a)pyrene	1.000	1.081	-8.1	119	-0.02
36	Dibenzo(a,h)anthracene	1.000	1.017	-1.7	114	-0.03
37	Benzo(g,h,i)perylene	1.000	1.029	-2.9	115	-0.04

(#) = Out of Range

SPCC's out = 0 CCC's out = 0