

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE070915\
 Data File : BE090076.D
 Acq On : 9 Jul 2015 16:56
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Jul 10 02:02:40 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE062915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 30 07:01:01 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	114	-0.02
2	1,4-Dioxane	1.000	0.856	14.4	102	-0.01
3	n-Nitrosodimethylamine	1.000	0.801	19.9	95	-0.01
4 S	2-Fluorophenol	1.000	0.974	2.6	112	-0.01
5 S	Phenol-d6	1.000	1.077	-7.7	125	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	124	-0.02
7 S	Nitrobenzene-d5	1.000	0.881	11.9	111	-0.02
8	Nitrobenzene	1.000	0.865	13.5	111	-0.01
9	Naphthalene	1.000	0.920	8.0	116	-0.01
10	Hexachlorobutadiene	1.000	0.936	6.4	118	-0.01
11	2-Methylnaphthalene	1.000	0.918	8.2	116	-0.01
12 I	Acenaphthene-d10	5.000	5.000	0.0	120	-0.01
13 S	2,4,6-Tribromophenol	1.000	0.877	12.3	108	0.00
14 S	2-Fluorobiphenyl	1.000	0.919	8.1	113	-0.01
15	Acenaphthylene	1.000	0.915	8.5	113	0.00
16	Acenaphthene	1.000	0.931	6.9	113	-0.01
17	Fluorene	1.000	0.892	10.8	110	-0.02
18 I	Phenanthrene-d10	5.000	5.000	0.0	107	-0.02
19	4-Bromophenyl-phenylether	1.000	0.914	8.6	100	-0.02
20	Hexachlorobenzene	1.000	0.914	8.6	102	-0.02
21	Pentachlorophenol	1.000	1.492	-49.2#	191	0.00
22	Phenanthrene	1.000	0.881	11.9	97	-0.02
23	Anthracene	1.000	0.891	10.9	97	-0.02
24	Fluoranthene	1.000	0.856	14.4	94	-0.01
25 I	Chrysene-d12	5.000	5.000	0.0	93	-0.01
26	Pyrene	1.000	0.997	0.3	94	-0.01
27 S	Terphenyl-d14	1.000	0.949	5.1	88	-0.01
28	Benzo(a)anthracene	1.000	0.930	7.0	88	-0.01
29	Chrysene	1.000	0.902	9.8	85	-0.01
30	Bis(2-ethylhexyl)phthalate	1.000	1.085	-8.5	121	-0.01
31	Indeno(1,2,3-cd)pyrene	1.000	0.962	3.8	91	-0.03
32 I	Perylene-d12	5.000	5.000	0.0	88	-0.02
33	Benzo(b)fluoranthene	1.000	1.045	-4.5	93	-0.02
34	Benzo(k)fluoranthene	1.000	1.002	-0.2	89	-0.02
35 C	Benzo(a)pyrene	1.000	1.036	-3.6	92	-0.02
36	Dibenzo(a,h)anthracene	1.000	0.985	1.5	89	-0.03
37	Benzo(g,h,i)perylene	1.000	0.992	0.8	89	-0.04

(#) = Out of Range

SPCC's out = 0 CCC's out = 0