

Data Path : V:\HPCHEM1\BNA E\DATA\BE070916\
 Data File : BE091994.D
 Acq On : 9 Jul 2016 12:09
 Operator : UM/SJ
 Sample : PB91976BSD
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 PB91976BSD

Quant Time: Jul 11 09:05:33 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE070516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 05 16:14:58 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.20	152	20167	5.00	ng	0.00
8) Naphthalene-d8	10.99	136	90561	5.00	ng	0.00
15) Acenaphthene-d10	14.86	164	56530	5.00	ng	0.00
24) Phenanthrene-d10	17.59	188	111819	5.00	ng	0.00
31) Chrysene-d12	21.76	240	195136	5.00	ng	0.00
40) Perylene-d12	24.17	264	134291	5.00	ng	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Fluorophenol	5.72	112	38812	8.04	ng	0.00
5) Phenol-d6	7.35	99	52404	7.82	ng	-0.02
9) Nitrobenzene-d5	9.37	82	28020	4.54	ng	-0.02
16) 2,4,6-Tribromophenol	16.35	330	12902	6.40	ng	0.00
17) 2-Fluorobiphenyl	13.48	172	63254	4.59	ng	0.00
34) Terphenyl-d14	20.22	244	86220	3.75	ng	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) 1,4-Dioxane	3.47	88	9058	3.31	ng	# 80
3) n-Nitrosodimethylamine	3.80	42	15016	4.25	ng	# 81
6) bis(2-Chloroethyl)ether	7.61	93	23851	4.08	ng	98
7) n-Nitroso-di-n-propylamine	9.00	70	19725	3.53	ng	# 100
10) Nitrobenzene	9.40	77	28407	4.24	ng	# 100
11) bis(2-Chloroethoxy)methane	10.40	93	30580	4.04	ng	# 12
12) Naphthalene	11.04	128	79477	3.77	ng	97
13) Hexachlorobutadiene	11.34	225	11534	3.60	ng	98
14) 2-Methylnaphthalene	12.66	142	52974	3.91	ng	96
18) Acenaphthylene	14.56	152	329585	4.64	ng	# 72
19) 2,6-Dinitrotoluene	14.44	165	62850	3.87	ng	# 10
20) Acenaphthene	14.92	154	56985	3.69	ng	# 1
21) 2,4-Dinitrotoluene	15.22	165	68804	3.82	ng	# 58
22) Fluorene	15.90	166	67246	4.12	ng	100
23) 4-Chlorophenyl-phenylether	15.90	204	31486	3.87	ng	100
25) 4-Bromophenyl-phenylether	16.78	248	18848	3.88	ng	97
26) Hexachlorobenzene	16.90	284	19208m	3.98	ng	
27) Pentachlorophenol	17.25	266	17027	6.64	ng	91
28) Phenanthrene	17.63	178	114434	3.93	ng	99
29) Anthracene	17.72	178	109299	4.01	ng	100
30) Fluoranthene	19.65	202	451084	3.60	ng	# 86
32) Benzidine	19.83	184	5263	0.55	ng	95
33) Pyrene	19.99	202	502199	3.53	ng	# 64
35) Benzo(a)anthracene	21.73	228	140020m	4.28	ng	
36) 3,3'-Dichlorobenzidine	21.67	252	14381	1.66	ng	# 61
37) Chrysene	21.78	228	186973m	4.04	ng	
38) Bis(2-ethylhexyl)phthalate	21.64	149	118861	3.14	ng	# 100
39) Indeno(1,2,3-cd)pyrene	26.68	276	666782	3.67	ng	98
41) Benzo(b)fluoranthene	23.42	252	142926	3.85	ng	98
42) Benzo(k)fluoranthene	23.48	252	153373	4.28	ng	97
43) Benzo(a)pyrene	24.06	252	139017	3.97	ng	96
44) Dibenzo(a,h)anthracene	26.70	278	138770	4.21	ng	96
45) Benzo(a,h,i)perylene	27.47	276	565665	4.21	ng	98

Data Path : V:\HPCHEM1\BNA E\DATA\BE070916\
Data File : BE091994.D
Acq On : 9 Jul 2016 12:09
Operator : UM/SJ
Sample : PB91976BSD
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
PB91976BSD

Quant Time: Jul 11 09:05:33 2016
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE070516.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 05 16:14:58 2016
Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
--------------------	------	------	----------	------	-------	----------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : V:\HPCHEM1\BNA E\DATA\BE070916\
Data File : BE091994.D
Acq On : 9 Jul 2016 12:09
Operator : UM/SJ
Sample : PB91976BSD
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
PB91976BSD

Quant Time: Jul 11 09:05:33 2016
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE070516.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 05 16:14:58 2016
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 8.20 min Scan# 323

Delta R.T. 0.00 min

Lab File: BE091994.D

Acq: 9 Jul 2016 12:09

Instrument :

BNA_E

ClientSampleId :

PB91976BSD




Tot Ion:152 Resp: 20167

Ion Ratio Lower Upper

152 100

150 153.3 123.9 185.9

115 58.5 48.3 72.5



#2

1,4-Dioxane

Concen: 3.31 ng

RT: 3.47 min Scan# 51

Delta R.T. 0.00 min

Lab File: BE091994.D

Acq: 9 Jul 2016 12:09




Tgt Ion: 88 Resp: 9058


Ion Ratio Lower Upper

88 100

58 94.7 63.0 94.4#

43 49.7 29.1 43.7#





#3

n-Nitrosodimethylamine

Concen: 4.25 ng

RT: 3.80 min Scan# 70

Delta R.T. -0.03 min

Lab File: BE091994.D

Acq: 9 Jul 2016 12:09

Instrument :

BNA_E

ClientSampleId :

PB91976BSD


Tot Ion: 42 Resp: 15016

Ion Ratio Lower Upper

42 100

74 95.7 93.4 140.0




44 8.8 5.0 7.6#



Abundance Ion 42.00 (41.70 to 42.70): BE0

Ion 74.00 (73.70 to 74.70): BE0

Ion 44.00 (43.70 to 44.70): BE0



#4

2-Fluorophenol

Concen: 8.04 ng

RT: 5.72 min Scan# 181

Delta R.T. 0.00 min

Lab File: BE091994.D

Acq: 9 Jul 2016 12:09


Tgt Ion: 112 Resp: 38812

Ion Ratio Lower Upper

112 100

64 66.9 53.7 80.5



63 36.3 29.5 44.3




Abundance Ion 112.00 (111.70 to 112.70): BE0

Ion 64.00 (63.70 to 64.70): BE0

Ion 63.00 (62.70 to 63.70): BE0





#5

Phenol-d6

Concen: 7.82 ng

RT: 7.35 min Scan# 274

Delta R.T. -0.02 min

Lab File: BE091994.D

Acq: 9 Jul 2016 12:09

Instrument :

BNA_E

ClientSampleId :

PB91976BSD


Tot Ion: 99 Resp: 52404

Ion Ratio Lower Upper




99 100

42 25.3 19.6 29.4

71 35.2 28.1 42.1



Abundance Ion 99.00 (98.70 to 99.70): BE0
Ion 42.00 (41.70 to 42.70): BE0
Ion 71.00 (70.70 to 71.70): BE0



#6

bis(2-Chloroethyl)ether

Concen: 4.08 ng

RT: 7.61 min Scan# 289

Delta R.T. -0.02 min

Lab File: BE091994.D

Acq: 9 Jul 2016 12:09


Tgt Ion: 93 Resp: 23851

Ion Ratio Lower Upper



93 100




63 85.0 66.2 99.4

95 33.0 25.5 38.3



Abundance Ion 93.00 (92.70 to 93.70): BE0
Ion 63.00 (62.70 to 63.70): BE0
Ion 95.00 (94.70 to 95.70): BE0





#7

n-Nitroso-di-n-propylamine

Concen: 3.53 ng

RT: 9.00 min Scan# 365

Delta R.T. -0.02 min

Lab File: BE091994.D

Acq: 9 Jul 2016 12:09

Instrument :

BNA_E

ClientSampleId :

PB91976BSD

Tot Ion: 70 Resp: 19725





Ion Ratio Lower Upper

70 100

42 73.2 58.4 87.6

101 8.4 6.4 9.6

130 16.6 0.0 0.0#



#8

Naphthalene-d8

Concen: 5.00 ng

RT: 10.99 min Scan# 456

Delta R.T. -0.00 min

Lab File: BE091994.D

Acq: 9 Jul 2016 12:09

Tgt Ion: 136 Resp: 90561


Ion Ratio Lower Upper


136 100

137 10.6 8.3 12.5

54 6.7 8.2 12.4#

68 5.1 5.9 8.9#





#9

Nitrobenzene-d5

Concen: 4.54 ng

RT: 9.37 min Scan# 384

Delta R.T. -0.02 min

Lab File: BE091994.D

Acq: 9 Jul 2016 12:09

Instrument :

BNA_E

ClientSampleId :

PB91976BSD


Tot Ion: 82 Resp: 28020

Ion Ratio Lower Upper




82 100

128 42.1 39.5 59.3

54 35.8 30.3 45.5



Abundance Ion 82.00 (81.70 to 82.70): BE0
Ion 128.00 (127.70 to 128.70): BE0
Ion 54.00 (53.70 to 54.70): BE0



#10

Nitrobenzene

Concen: 4.24 ng

RT: 9.40 min Scan# 386

Delta R.T. -0.02 min

Lab File: BE091994.D

Acq: 9 Jul 2016 12:09


Tgt Ion: 77 Resp: 28407

Ion Ratio Lower Upper



77 100


123 38.4 0.0 0.0#

65 13.8 11.1 16.7



Abundance Ion 77.00 (76.70 to 77.70): BE0
Ion 123.00 (122.70 to 123.70): BE0
Ion 65.00 (64.70 to 65.70): BE0





#11

bis(2-Chloroethoxy)methane

Concen: 4.04 ng

RT: 10.40 min Scan# 432

Delta R.T. -0.05 min

Lab File: BE091994.D

Acq: 9 Jul 2016 12:09

Instrument :

BNA_E

ClientSampleId :

PB91976BSD





Tot Ion: 93 Resp: 30580

Ion Ratio Lower Upper

93 100

95 31.4 57.6 86.4#

123 0.0 100.2 150.4#



#12

Naphthalene

Concen: 3.77 ng

RT: 11.04 min Scan# 458

Delta R.T. -0.00 min

Lab File: BE091994.D

Acq: 9 Jul 2016 12:09




Tgt Ion: 128 Resp: 79477


Ion Ratio Lower Upper

128 100

129 10.9 10.4 15.6

127 13.0 10.2 15.2









#13
Hexachlorobutadiene
Concen: 3.60 ng
RT: 11.34 min Scan# 470
Delta R.T. -0.00 min
Lab File: BE091994.D
Acq: 9 Jul 2016 12:09




Instrument :
BNA_E
ClientSampleId :
PB91976BSD


Tot Ion:225 Resp: 11534
Ion Ratio Lower Upper
225 100
223 60.7 49.8 74.8
227 65.3 50.6 75.8



#14
2-Methylnaphthalene
Concen: 3.91 ng
RT: 12.66 min Scan# 551
Delta R.T. -0.02 min
Lab File: BE091994.D
Acq: 9 Jul 2016 12:09

Tgt Ion:142 Resp: 52974
Ion Ratio Lower Upper
142 100
141 93.6 71.9 107.9
115 39.9 29.9 44.9





#15

Acenaphthene-d10

Concen: 5.00 ng

RT: 14.86 min Scan# 694

Delta R.T. 0.00 min

Lab File: BE091994.D

Acq: 9 Jul 2016 12:09

Instrument :

BNA_E

ClientSampleId :

PB91976BSD


Tot Ion:164 Resp: 56530

Ion Ratio Lower Upper

164 100

162 96.6 71.8 107.8




160 40.1 28.0 42.0



Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E



#16

2,4,6-Tribromophenol

Concen: 6.40 ng

RT: 16.35 min Scan# 762

Delta R.T. 0.00 min

Lab File: BE091994.D

Acq: 9 Jul 2016 12:09


Tgt Ion:330 Resp: 12902

Ion Ratio Lower Upper

330 100

332 97.3 77.0 115.6



141 173.8 138.4 207.6




Abundance Ion 329.80 (329.50 to 330.50): E

Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E









#17
2-Fluorobiphenyl
Concen: 4.59 ng
RT: 13.48 min Scan# 622
Delta R.T. 0.00 min
Lab File: BE091994.D
Acq: 9 Jul 2016 12:09




Instrument :
BNA_E
ClientSampleId :
PB91976BSD


Tot Ion:172 Resp: 63254
Ion Ratio Lower Upper
172 100
171 30.3 24.3 36.5
170 17.8 14.9 22.3



#18
Acenaphthylene
Concen: 4.64 ng
RT: 14.56 min Scan# 684
Delta R.T. 0.00 min
Lab File: BE091994.D
Acq: 9 Jul 2016 12:09

Tgt Ion:152 Resp: 329585
Ion Ratio Lower Upper
152 100
151 5.6 19.4 29.2#
153 11.1 14.4 21.6#









#19
 2,6-Dinitrotoluene
 Concen: 3.87 ng
 RT: 14.44 min Scan# 680
 Delta R.T. 0.00 min
 Lab File: BE091994.D
 Acq: 9 Jul 2016 12:09




Instrument :
 BNA_E
 ClientSampleId :
 PB91976BSD


Tot Ion:165 Resp: 62850
 Ion Ratio Lower Upper
 165 100
 63 0.0 65.9 98.9#
 89 0.0 59.1 88.7#



#20
 Acenaphthene
 Concen: 3.69 ng
 RT: 14.92 min Scan# 696
 Delta R.T. 0.00 min
 Lab File: BE091994.D
 Acq: 9 Jul 2016 12:09

Tgt Ion:154 Resp: 56985
 Ion Ratio Lower Upper
 154 100
 153 370.8 88.1 132.1#
 152 150.8 44.2 66.2#





#21

2,4-Dinitrotoluene

Concen: 3.82 ng

RT: 15.22 min Scan# 706

Delta R.T. -0.03 min

Lab File: BE091994.D

Acq: 9 Jul 2016 12:09

Instrument :

BNA_E

ClientSampleId :

PB91976BSD

Tot Ion:165 Resp: 68804


Ion Ratio Lower Upper

165 100

63 68.0 0.0 0.0#

89 76.6 99.8 149.8#

182 0.5 0.0 0.0#




Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BE0


Ion 89.00 (88.70 to 89.70): BE0

Ion 182.00 (181.70 to 182.70): E

Time-->



Time-->



#22

Fluorene

Concen: 4.12 ng

RT: 15.90 min Scan# 736

Delta R.T. 0.00 min

Lab File: BE091994.D

Acq: 9 Jul 2016 12:09


Tgt Ion:166 Resp: 67246

Ion Ratio Lower Upper

166 100

165 98.3 78.6 117.8

167 13.3 10.8 16.2




Abundance Ion 166.00 (165.70 to 166.70): E


Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E

Time-->







Time-->



#23
 4-Chlorophenyl-phenylether
 Concen: 3.87 ng
 RT: 15.90 min Scan# 736
 Delta R.T. 0.00 min
 Lab File: BE091994.D
 Acq: 9 Jul 2016 12:09




Instrument :
 BNA_E
 ClientSampleID :
 PB91976BSD


Tot Ion:204 Resp: 31486
 Ion Ratio Lower Upper
 204 100
 206 33.9 27.8 41.6
 141 247.3 197.6 296.4



#24
 Phenanthrene-d10
 Concen: 5.00 ng
 RT: 17.59 min Scan# 835
 Delta R.T. 0.00 min
 Lab File: BE091994.D
 Acq: 9 Jul 2016 12:09

Tgt Ion:188 Resp: 111819
 Ion Ratio Lower Upper
 188 100
 94 8.8 4.1 6.1#
 80 8.3 3.4 5.2#





#25

4-Bromophenyl-phenylether

Concen: 3.88 ng

RT: 16.78 min Scan# 785

Delta R.T. 0.00 min

Lab File: BE091994.D

Acq: 9 Jul 2016 12:09

Instrument :

BNA_E

ClientSampleId :

PB91976BSD


Tot Ion:248 Resp: 18848

Ion Ratio Lower Upper




248 100

250 97.7 75.7 113.5

141 83.0 64.6 97.0



Abundance Ion 248.00 (247.70 to 248.70): E
Ion 250.00 (249.70 to 250.70): E
Ion 141.00 (140.70 to 141.70): E



#26

Hexachlorobenzene

Concen: 3.98 ng m

RT: 16.90 min Scan# 792

Delta R.T. 0.00 min

Lab File: BE091994.D

Acq: 9 Jul 2016 12:09


Tgt Ion:284 Resp: 19208

Ion Ratio Lower Upper



284 100


142 40.3 31.4 47.2

249 31.2 25.3 37.9



Abundance Ion 283.80 (283.50 to 284.50): E
Ion 142.00 (141.70 to 142.70): E
Ion 249.00 (248.70 to 249.70): E









#27
 Pentachlorophenol
 Concen: 6.64 ng
 RT: 17.25 min Scan# 812
 Delta R.T. -0.02 min
 Lab File: BE091994.D
 Acq: 9 Jul 2016 12:09




Instrument :
 BNA_E
 ClientSampleId :
 PB91976BSD


Tot Ion:266 Resp: 17027
 Ion Ratio Lower Upper
 266 100
 268 63.5 60.5 90.7
 264 60.9 51.0 76.4



#28
 Phenanthrene
 Concen: 3.93 ng
 RT: 17.63 min Scan# 838
 Delta R.T. -0.01 min
 Lab File: BE091994.D
 Acq: 9 Jul 2016 12:09

Tgt Ion:178 Resp: 114434
 Ion Ratio Lower Upper
 178 100
 176 19.5 15.4 23.2
 179 15.4 12.9 19.3





#29

Anthracene

Concen: 4.01 ng

RT: 17.72 min Scan# 846

Delta R.T. -0.01 min

Lab File: BE091994.D

Acq: 9 Jul 2016 12:09

Instrument :

BNA_E

ClientSampleId :

PB91976BSD


Tot Ion:178 Resp: 109299

Ion Ratio Lower Upper

178 100

176 18.9 15.1 22.7




179 15.2 12.1 18.1



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#30

Fluoranthene

Concen: 3.60 ng

RT: 19.65 min Scan# 984

Delta R.T. -0.00 min

Lab File: BE091994.D

Acq: 9 Jul 2016 12:09


Tgt Ion:202 Resp: 451084

Ion Ratio Lower Upper

202 100

101 0.0 9.4 14.0#



203 19.6 14.3 21.5




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E





#31

Chrysene-d12

Concen: 5.00 ng

RT: 21.76 min Scan# 1064

Delta R.T. 0.00 min

Lab File: BE091994.D

Acq: 9 Jul 2016 12:09

Instrument :

BNA_E

ClientSampleId :

PB91976BSD


Tot Ion:240 Resp: 195136

Ion Ratio Lower Upper

240 100

120 1.4 1.2 1.8




236 26.7 19.8 29.8



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#32

Benzidine

Concen: 0.55 ng

RT: 19.83 min Scan# 992

Delta R.T. -0.00 min

Lab File: BE091994.D

Acq: 9 Jul 2016 12:09


Tgt Ion:184 Resp: 5263

Ion Ratio Lower Upper

184 100

185 16.2 11.4 17.2



183 12.5 11.8 17.6




Abundance Ion 184.00 (183.70 to 184.70): E

Ion 185.00 (184.70 to 185.70): E

Ion 183.00 (182.70 to 183.70): E





#33

Pvrene

Concen: 3.53 ng

RT: 19.99 min Scan# 999

Delta R.T. -0.02 min

Lab File: BE091994.D

Acq: 9 Jul 2016 12:09

Instrument :

BNA_E

ClientSampleId :

PB91976BSD


Tot Ion:202 Resp: 502199

Ion Ratio Lower Upper

202 100

200 5.6 16.9 25.3#




203 0.0 14.1 21.1#



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E



#34

Terphenyl-d14

Concen: 3.75 ng

RT: 20.22 min Scan# 1009

Delta R.T. -0.00 min

Lab File: BE091994.D

Acq: 9 Jul 2016 12:09


Tgt Ion:244 Resp: 86220

Ion Ratio Lower Upper

244 100

212 6.8 6.5 9.7



122 0.5 2.9 4.3#




Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E





#35

Benzo(a)anthracene

Concen: 4.28 ng/mL

RT: 21.73 min Scan# 1063

Delta R.T. 0.00 min

Lab File: BE091994.D

Acq: 9 Jul 2016 12:09

Instrument :

BNA_E

ClientSampleId :

PB91976BSD


Tot Ion:228 Resp: 140020

Ion Ratio Lower Upper

228 100

226 33.7 21.0 31.4#




229 15.6 15.8 23.8#



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#36

3,3'-Dichlorobenzidine

Concen: 1.66 ng/mL

RT: 21.67 min Scan# 1061

Delta R.T. -0.03 min

Lab File: BE091994.D

Acq: 9 Jul 2016 12:09


Tgt Ion:252 Resp: 14381

Ion Ratio Lower Upper

252 100

254 44.4 61.5 92.3#



126 29.3 2.5 3.7#




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 254.00 (253.70 to 254.70): E

Ion 126.00 (125.70 to 126.70): E





#37

Chrysene

Concen: 4.04 ng/mL

RT: 21.78 min Scan# 1065

Delta R.T. 0.00 min

Lab File: BE091994.D

Acq: 9 Jul 2016 12:09

Instrument :

BNA_E

ClientSampleID :

PB91976BSD


Tot Ion: 228 Resp: 186973

Ion Ratio Lower Upper

228 100

226 32.4 27.0 40.6




229 17.8 13.8 20.8



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#38

Bis(2-ethylhexyl)phthalate

Concen: 3.14 ng/mL

RT: 21.64 min Scan# 1060

Delta R.T. -0.03 min

Lab File: BE091994.D

Acq: 9 Jul 2016 12:09


Tgt Ion: 149 Resp: 118861

Ion Ratio Lower Upper

149 100

167 0.0 0.0 0.0



279 0.0 0.0 0.0




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E









#39
 Indeno(1,2,3-cd)pyrene
 Concen: 3.67 ng
 RT: 26.68 min Scan# 1401
 Delta R.T. -0.02 min
 Lab File: BE091994.D
 Acq: 9 Jul 2016 12:09

Instrument :
 BNA_E
 ClientSampleId :
 PB91976BSD

Tot Ion:276 Resp: 666782
 Ion Ratio Lower Upper
 276 100
 138 26.4 19.7 29.5
 277 24.8 20.1 30.1




Abundance Ion 276.00 (275.70 to 276.70): E
 Ion 138.00 (137.70 to 138.70): E
 Ion 277.00 (276.70 to 277.70): E






#40
 Perylene-d12
 Concen: 5.00 ng
 RT: 24.17 min Scan# 1232
 Delta R.T. 0.00 min
 Lab File: BE091994.D
 Acq: 9 Jul 2016 12:09

Tgt Ion:264 Resp: 134291
 Ion Ratio Lower Upper
 264 100
 260 23.7 20.3 30.5
 265 22.7 17.6 26.4



Abundance Ion 264.00 (263.70 to 264.70): E
 Ion 260.00 (259.70 to 260.70): E
 Ion 265.00 (264.70 to 265.70): E





#41

Benzo(b)fluoranthene

Concen: 3.85 ng

RT: 23.42 min Scan# 1167

Delta R.T. -0.01 min

Lab File: BE091994.D

Acq: 9 Jul 2016 12:09

Instrument :

BNA_E

ClientSampleId :

PB91976BSD


Tot Ion:252 Resp: 142926

Ion Ratio Lower Upper




252 100

253 21.1 17.8 26.6

125 10.6 9.4 14.2



Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E



#42

Benzo(k)fluoranthene

Concen: 4.28 ng

RT: 23.48 min Scan# 1172

Delta R.T. 0.00 min

Lab File: BE091994.D

Acq: 9 Jul 2016 12:09


Tgt Ion:252 Resp: 153373

Ion Ratio Lower Upper



252 100


253 23.0 19.4 29.2

125 9.2 8.3 12.5



Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E





#43

Benzo(a)pyrene

Concen: 3.97 ng

RT: 24.06 min Scan# 1222

Delta R.T. 0.00 min

Lab File: BE091994.D

Acq: 9 Jul 2016 12:09

Instrument :

BNA_E

ClientSampleId :

PB91976BSD


Tot Ion:252 Resp: 139017

Ion Ratio Lower Upper




252 100

253 22.2 18.9 28.3

125 10.9 10.5 15.7



Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E



#44

Dibenzo(a,h)anthracene

Concen: 4.21 ng

RT: 26.70 min Scan# 1402

Delta R.T. -0.02 min

Lab File: BE091994.D

Acq: 9 Jul 2016 12:09


Tgt Ion:278 Resp: 138770

Ion Ratio Lower Upper



278 100


139 17.9 12.6 18.8

279 23.0 20.0 30.0



Abundance Ion 278.00 (277.70 to 278.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 279.00 (278.70 to 279.70): E





#45

Benzo(a,h,i)perylene

Concen: 4.21 ng

RT: 27.47 min Scan# 1447

Delta R.T. 0.00 min

Lab File: BE091994.D


Acq: 9 Jul 2016 12:09

Instrument :

BNA_E

ClientSampleId :

PB91976BSD




Tot Ion:276 Resp: 565665

Ion Ratio Lower Upper

276 100

277 24.6 19.5 29.3

138 20.3 17.7 26.5



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

