

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE071015\
 Data File : BE090081.D
 Acq On : 10 Jul 2015 12:37
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Jul 10 23:08:10 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE062915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 30 07:01:01 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	120	-0.02
2	1,4-Dioxane	0.686	0.613	10.6	111	-0.01
3	n-Nitrosodimethylamine	0.930	0.758	18.5	102	-0.02
4 S	2-Fluorophenol	1.149	1.057	8.0	111	-0.01
5 S	Phenol-d6	1.608	1.524	5.2	116	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	120	-0.02
7 S	Nitrobenzene-d5	0.397	0.348	12.3	107	-0.02
8	Nitrobenzene	0.421	0.366	13.1	108	-0.02
9	Naphthalene	1.130	1.043	7.7	113	-0.01
10	Hexachlorobutadiene	0.178	0.170	4.5	117	-0.01
11	2-Methylnaphthalene	0.755	0.656	13.1	106	-0.02
12 I	Acenaphthene-d10	1.000	1.000	0.0	109	-0.01
13 S	2,4,6-Tribromophenol	0.147	0.123	16.3	93	0.00
14 S	2-Fluorobiphenyl	1.503	1.381	8.1	103	-0.02
15	Acenaphthylene	9.019	8.281	8.2	103	-0.01
16	Acenaphthene	1.300	1.208	7.1	102	-0.01
17	Fluorene	1.768	1.582	10.5	100	-0.02
18 I	Phenanthrene-d10	1.000	1.000	0.0	96	-0.02
19	4-Bromophenyl-phenylether	0.213	0.193	9.4	89	-0.02
20	Hexachlorobenzene	0.885	0.797	9.9	90	-0.02
21	Pentachlorophenol	0.086	0.445	-417.4#	526#	0.00
22	Phenanthrene	1.281	1.118	12.7	86	-0.02
23	Anthracene	1.172	1.046	10.8	87	0.00
24	Fluoranthene	1.431	1.226	14.3	84	-0.02
25 I	Chrysene-d12	1.000	1.000	0.0	84	-0.01
26	Pyrene	1.192	1.182	0.8	85	-0.02
27 S	Terphenyl-d14	0.833	0.782	6.1	79	-0.01
28	Benzo(a)anthracene	1.249	1.153	7.7	79	-0.01
29	Chrysene	1.297	1.169	9.9	77	-0.01
30	Bis(2-ethylhexyl)phthalate	0.398	0.419	-5.3	96	-0.02
31	Indeno(1,2,3-cd)pyrene	1.243	1.172	5.7	81	-0.03
32 I	Perylene-d12	1.000	1.000	0.0	79	-0.02
33	Benzo(b)fluoranthene	1.103	1.113	-0.9	81	-0.02
34	Benzo(k)fluoranthene	1.232	1.250	-1.5	81	-0.02
35 C	Benzo(a)pyrene	1.019	1.035	-1.6	81	-0.02
36	Dibenzo(a,h)anthracene	1.046	1.013	3.2	78	-0.03
37	Benzo(g,h,i)perylene	1.060	1.037	2.2	79	-0.04

(#) = Out of Range

SPCC's out = 0 CCC's out = 0