

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE071015\
 Data File : BE090081.D
 Acq On : 10 Jul 2015 12:37
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Jul 10 23:08:10 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE062915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 30 07:01:01 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	120	-0.02
2	1,4-Dioxane	1.000	0.893	10.7	111	-0.01
3	n-Nitrosodimethylamine	1.000	0.815	18.5	102	-0.02
4 S	2-Fluorophenol	1.000	0.920	8.0	111	-0.01
5 S	Phenol-d6	1.000	0.948	5.2	116	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	120	-0.02
7 S	Nitrobenzene-d5	1.000	0.876	12.4	107	-0.02
8	Nitrobenzene	1.000	0.869	13.1	108	-0.02
9	Naphthalene	1.000	0.923	7.7	113	-0.01
10	Hexachlorobutadiene	1.000	0.956	4.4	117	-0.01
11	2-Methylnaphthalene	1.000	0.869	13.1	106	-0.02
12 I	Acenaphthene-d10	5.000	5.000	0.0	109	-0.01
13 S	2,4,6-Tribromophenol	1.000	0.837	16.3	93	0.00
14 S	2-Fluorobiphenyl	1.000	0.919	8.1	103	-0.02
15	Acenaphthylene	1.000	0.918	8.2	103	-0.01
16	Acenaphthene	1.000	0.929	7.1	102	-0.01
17	Fluorene	1.000	0.895	10.5	100	-0.02
18 I	Phenanthrene-d10	5.000	5.000	0.0	96	-0.02
19	4-Bromophenyl-phenylether	1.000	0.909	9.1	89	-0.02
20	Hexachlorobenzene	1.000	0.901	9.9	90	-0.02
21	Pentachlorophenol	1.000	4.060	-306.0#	526	0.00
22	Phenanthrene	1.000	0.872	12.8	86	-0.02
23	Anthracene	1.000	0.892	10.8	87	0.00
24	Fluoranthene	1.000	0.857	14.3	84	-0.02
25 I	Chrysene-d12	5.000	5.000	0.0	84	-0.01
26	Pyrene	1.000	0.992	0.8	85	-0.02
27 S	Terphenyl-d14	1.000	0.938	6.2	79	-0.01
28	Benzo(a)anthracene	1.000	0.923	7.7	79	-0.01
29	Chrysene	1.000	0.902	9.8	77	-0.01
30	Bis(2-ethylhexyl)phthalate	1.000	0.992	0.8	96	-0.02
31	Indeno(1,2,3-cd)pyrene	1.000	0.942	5.8	81	-0.03
32 I	Perylene-d12	5.000	5.000	0.0	79	-0.02
33	Benzo(b)fluoranthene	1.000	1.009	-0.9	81	-0.02
34	Benzo(k)fluoranthene	1.000	1.014	-1.4	81	-0.02
35 C	Benzo(a)pyrene	1.000	1.016	-1.6	81	-0.02
36	Dibenzo(a,h)anthracene	1.000	0.969	3.1	78	-0.03
37	Benzo(g,h,i)perylene	1.000	0.979	2.1	79	-0.04

(#) = Out of Range

SPCC's out = 0 CCC's out = 0