

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE071015\
 Data File : BE090083.D
 Acq On : 10 Jul 2015 15:04
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Jul 10 23:13:55 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE062915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 30 07:01:01 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	120	-0.01
2	1,4-Dioxane	0.686	0.542	21.0	99	-0.01
3	n-Nitrosodimethylamine	0.930	0.731	21.4	98	-0.01
4 S	2-Fluorophenol	1.149	1.282	-11.6	135	0.00
5 S	Phenol-d6	1.608	2.349	-46.1#	179#	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	127	-0.01
7 S	Nitrobenzene-d5	0.397	0.414	-4.3	135	0.00
8	Nitrobenzene	0.421	0.339	19.5	106	0.00
9	Naphthalene	1.130	0.985	12.8	113	-0.01
10	Hexachlorobutadiene	0.178	0.160	10.1	117	0.00
11	2-Methylnaphthalene	0.755	0.626	17.1	107	-0.01
12 I	Acenaphthene-d10	1.000	1.000	0.0	116	0.00
13 S	2,4,6-Tribromophenol	0.147	0.202	-37.4#	163#	0.00
14 S	2-Fluorobiphenyl	1.503	1.846	-22.8	146	-0.01
15	Acenaphthylene	9.019	7.740	14.2	102	0.00
16	Acenaphthene	1.300	1.126	13.4	102	-0.01
17	Fluorene	1.768	1.461	17.4	98	-0.02
18 I	Phenanthrene-d10	1.000	1.000	0.0	101	-0.02
19	4-Bromophenyl-phenylether	0.213	0.180	15.5	88	-0.02
20	Hexachlorobenzene	0.885	0.749	15.4	89	-0.02
21	Pentachlorophenol	0.086	0.065	24.4	81	0.00
22	Phenanthrene	1.281	1.069	16.5	87	-0.02
23	Anthracene	1.172	0.973	17.0	86	0.00
24	Fluoranthene	1.431	1.138	20.5	82	-0.02
25 I	Chrysene-d12	1.000	1.000	0.0	88	-0.01
26	Pyrene	1.192	1.092	8.4	82	-0.01
27 S	Terphenyl-d14	0.833	1.098	-31.8#	116	-0.01
28	Benzo(a)anthracene	1.249	1.072	14.2	77	-0.01
29	Chrysene	1.297	1.084	16.4	75	-0.01
30	Bis(2-ethylhexyl)phthalate	0.398	0.366	8.0	88	-0.02
31	Indeno(1,2,3-cd)pyrene	1.243	1.044	16.0	75	-0.03
32 I	Perylene-d12	1.000	1.000	0.0	82	-0.02
33	Benzo(b)fluoranthene	1.103	1.020	7.5	77	-0.02
34	Benzo(k)fluoranthene	1.232	1.154	6.3	77	-0.02
35 C	Benzo(a)pyrene	1.019	0.940	7.8	76	-0.02
36	Dibenzo(a,h)anthracene	1.046	0.919	12.1	73	-0.03
37	Benzo(g,h,i)perylene	1.060	0.935	11.8	73	-0.04

(#) = Out of Range

SPCC's out = 0 CCC's out = 0