

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE071118\
 Data File : BE096892.D
 Acq On : 11 Jul 2018 16:31
 Operator : SJ/JU
 Sample : J3779-01
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 F1H25

Manual Integrations
 APPROVED

Sohil
 7/12/2018 11:20:34 AM

Quant Time: Jul 11 18:20:24 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE070618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 10 02:37:30 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.40	152	2057	0.40	ng/ul	0.00
2) Naphthalene-d8	10.14	136	11909	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.03	164	7580	0.40	ng/ul	0.00
10) Phenanthrene-d10	16.78	188	16653m	0.40	ng/ul	0.00
16) Chrysene-d12	20.98	240	13132m	0.40	ng/ul	0.00
20) Perylene-d12	23.22	264	14117m	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	11.74	152	8664	0.44	ng/ul	0.00
14) Fluoranthene-d10	18.81	212	19247m	0.46	ng/ul	0.00
Target Compounds						
3) Naphthalene	10.19	128	1721	0.060	ng/ul#	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE071118\
Data File : BE096892.D
Acq On : 11 Jul 2018 16:31
Operator : SJ/JU
Sample : J3779-01
Misc :
ALS Vial : 3 Sample Multiplier: 1




Instrument :
BNA_E
ClientSampleId :
F1H25

Quant Time: Jul 11 18:20:24 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE070618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 10 02:37:30 2018
Response via : Initial Calibration

Manual Integrations
APPROVED

Sohil
7/12/2018 11:20:34 AM





#3

Naphthalene

Concen: 0.060 ng/ul

RT: 10.19 min Scan# 1483

Delta R.T. -0.00 min

Lab File: BE096892.D

Acq: 11 Jul 2018 16:31

Instrument :

BNA_E

ClientSampleId :

F1H25

Tgt Ion:	128	Resp:	1721
Ion Ratio	Lower	Upper	
128	100		
129	15.1	11.3	16.9
127	0.0	4.0	6.0

Manual Integrations
APPROVED

Sohil
7/12/2018 11:20:34 AM

