

Data Path : Z:\HPCHEM1\BNA_E\Data\BE071317\
 Data File : BE093450.D
 Acq On : 13 Jul 2017 9:00
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 13 15:26:50 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE070217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 03 06:40:23 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	132	-0.01
2	1,4-Dioxane	0.400	0.327	18.3	108	-0.01
3	n-Nitrosodimethylamine	0.400	0.411	-2.7	166	0.00
4 S	2-Fluorophenol	0.400	0.374	6.5	127	-0.01
5 S	Phenol-d6	0.400	0.403	-0.8	142	0.00
6	bis(2-Chloroethyl)ether	0.400	0.392	2.0	133	-0.01
7 I	Naphthalene-d8	0.400	0.400	0.0	146	0.00
8 S	Nitrobenzene-d5	0.400	0.427	-6.7	168	0.00
9	Naphthalene	0.400	0.385	3.8	144	0.00
10	Hexachlorobutadiene	0.400	0.381	4.8	141	-0.02
11 SURR	2-Methylnaphthalene-d10	0.400	0.405	-1.3	153	0.00
12	2-Methylnaphthalene	0.400	0.404	-1.0	153	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	149	0.00
14 S	2,4,6-Tribromophenol	0.400	0.278	30.5#	114	0.00
15 S	2-Fluorobiphenyl	0.400	0.384	4.0	149	0.00
16	Acenaphthylene	0.400	0.384	4.0	151	0.00
17	Acenaphthene	0.400	0.386	3.5	150	0.00
18	Fluorene	0.400	0.359	10.3	139	-0.02
19 I	Phenanthrene-d10	0.400	0.400	0.0	98	0.00
20	4-Bromophenyl-phenylether	0.400	0.553	-38.3#	147	0.00
21	Hexachlorobenzene	0.400	0.496	-24.0	136	0.00
22	Pentachlorophenol	0.400	0.596	-49.0#	157	0.00
23	Phenanthrene	0.400	0.421	-5.2	113	0.00
24	Anthracene	0.400	0.389	2.8	101	0.00
25 SURR	Fluoranthene-d10	0.400	0.439	-9.7	118	0.00
26	Fluoranthene	0.400	0.439	-9.7	118	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	109	-0.01
28	Pyrene	0.400	0.420	-5.0	117	0.00
29 S	Terphenyl-d14	0.400	0.409	-2.2	113	0.00
30	Benzo(a)anthracene	0.400	0.399	0.3	114	0.00
31	Chrysene	0.400	0.400	0.0	110	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.429	-7.2	142	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.426	-6.5	119	-0.01
34 I	Perylene-d12	0.400	0.400	0.0	117	0.00
35	Benzo(b)fluoranthene	0.400	0.377	5.8	113	0.00
36	Benzo(k)fluoranthene	0.400	0.371	7.3	114	-0.01
37 C	Benzo(a)pyrene	0.400	0.388	3.0	120	0.00
38	Dibenzo(a,h)anthracene	0.400	0.385	3.8	117	-0.02
39	Benzo(g,h,i)perylene	0.400	0.383	4.3	118	-0.01

Data Path : Z:\HPCHEM1\BNA_E\Data\BE071317\
Data File : BE093450.D
Acq On : 13 Jul 2017 9:00
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 13 15:26:50 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE070217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 03 06:40:23 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0		CCC's out = 0		