

Data Path : Z:\HPCHEM1\BNA_E\Data\BE071317\
 Data File : BE093450.D
 Acq On : 13 Jul 2017 9:00
 Operator : SJ/JU
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDC00.4

Quant Time: Jul 13 15:26:50 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE070217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 03 06:40:23 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	132	-0.01
2	1,4-Dioxane	0.847	0.693	18.2	108	-0.01
3	n-Nitrosodimethylamine	0.444	0.444	0.0	166#	0.00
4 S	2-Fluorophenol	1.207	1.130	6.4	127	-0.01
5 S	Phenol-d6	1.780	1.795	-0.8	142	0.00
6	bis(2-Chloroethyl)ether	1.339	1.314	1.9	133	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	146	0.00
8 S	Nitrobenzene-d5	0.256	0.273	-6.6	168#	0.00
9	Naphthalene	4.190	4.037	3.7	144	0.00
10	Hexachlorobutadiene	0.159	0.151	5.0	141	-0.02
11 SURR	2-Methylnaphthalene-d10	0.638	0.646	-1.3	153#	0.00
12	2-Methylnaphthalene	0.704	0.710	-0.9	153#	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	149	0.00
14 S	2,4,6-Tribromophenol	0.126	0.087	31.0#	114	0.00
15 S	2-Fluorobiphenyl	1.553	1.491	4.0	149	0.00
16	Acenaphthylene	1.666	1.598	4.1	151#	0.00
17	Acenaphthene	1.220	1.178	3.4	150#	0.00
18	Fluorene	1.675	1.502	10.3	139	-0.02
19 I	Phenanthrene-d10	1.000	1.000	0.0	98	0.00
20	4-Bromophenyl-phenylether	0.219	0.303	-38.4#	147	0.00
21	Hexachlorobenzene	0.237	0.294	-24.1	136	0.00
22	Pentachlorophenol	0.035	0.046	-31.4#	157#	0.00
23	Phenanthrene	1.454	1.529	-5.2	113	0.00
24	Anthracene	0.891	0.868	2.6	101	0.00
25 SURR	Fluoranthene-d10	4.823	5.289	-9.7	118	0.00
26	Fluoranthene	1.367	1.498	-9.6	118	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	109	-0.01
28	Pyrene	1.124	1.182	-5.2	117	0.00
29 S	Terphenyl-d14	0.714	0.731	-2.4	113	0.00
30	Benzo(a)anthracene	1.080	1.077	0.3	114	0.00
31	Chrysene	1.130	1.130	0.0	110	0.00
32	Bis(2-ethylhexyl)phthalate	1.555	1.668	-7.3	142	-0.01
33	Indeno(1,2,3-cd)pyrene	0.992	1.055	-6.4	119	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	117	0.00
35	Benzo(b)fluoranthene	1.292	1.217	5.8	113	0.00
36	Benzo(k)fluoranthene	1.278	1.187	7.1	114	-0.01
37 C	Benzo(a)pyrene	1.132	1.098	3.0	120	0.00
38	Dibenzo(a,h)anthracene	1.099	1.058	3.7	117	-0.02
39	Benzo(g,h,i)perylene	1.110	1.061	4.4	118	-0.01

Data Path : Z:\HPCHEM1\BNA_E\Data\BE071317\
Data File : BE093450.D
Acq On : 13 Jul 2017 9:00
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 13 15:26:50 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE070217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 03 06:40:23 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		