

Data Path : Z:\HPCHEM1\BNA_E\Data\BE071317\
 Data File : BE093461.D
 Acq On : 13 Jul 2017 16:41
 Operator : SJ/JU
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDC00.4

Quant Time: Jul 13 17:18:19 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE070217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 03 06:40:23 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	139	0.00
2	1,4-Dioxane	0.847	0.780	7.9	128	0.00
3	n-Nitrosodimethylamine	0.444	0.464	-4.5	183#	0.00
4 S	2-Fluorophenol	1.207	1.115	7.6	132	0.00
5 S	Phenol-d6	1.780	1.760	1.1	147	0.00
6	bis(2-Chloroethyl)ether	1.339	1.324	1.1	141	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	156#	0.00
8 S	Nitrobenzene-d5	0.256	0.281	-9.8	186#	0.00
9	Naphthalene	4.190	4.020	4.1	154#	0.00
10	Hexachlorobutadiene	0.159	0.151	5.0	151#	-0.02
11 SURR	2-Methylnaphthalene-d10	0.638	0.635	0.5	161#	0.00
12	2-Methylnaphthalene	0.704	0.698	0.9	161#	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	153#	0.00
14 S	2,4,6-Tribromophenol	0.126	0.085	32.5#	113	0.00
15 S	2-Fluorobiphenyl	1.553	1.537	1.0	157#	0.00
16	Acenaphthylene	1.666	1.585	4.9	154#	0.00
17	Acenaphthene	1.220	1.172	3.9	153#	0.00
18	Fluorene	1.675	1.504	10.2	143	-0.02
19 I	Phenanthrene-d10	1.000	1.000	0.0	97	0.00
20	4-Bromophenyl-phenylether	0.219	0.320	-46.1#	153#	0.00
21	Hexachlorobenzene	0.237	0.300	-26.6#	137	0.00
22	Pentachlorophenol	0.035	0.046	-31.4#	157#	0.00
23	Phenanthrene	1.454	1.525	-4.9	111	0.00
24	Anthracene	0.891	0.883	0.9	102	0.00
25 SURR	Fluoranthene-d10	4.823	5.199	-7.8	114	0.00
26	Fluoranthene	1.367	1.471	-7.6	115	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	105	-0.01
28	Pyrene	1.124	1.179	-4.9	113	0.00
29 S	Terphenyl-d14	0.714	0.741	-3.8	111	0.00
30	Benzo(a)anthracene	1.080	1.040	3.7	106	0.00
31	Chrysene	1.130	1.127	0.3	106	0.00
32	Bis(2-ethylhexyl)phthalate	1.555	1.444	7.1	119	-0.01
33	Indeno(1,2,3-cd)pyrene	0.992	1.051	-5.9	114	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	110	0.00
35	Benzo(b)fluoranthene	1.292	1.238	4.2	108	0.00
36	Benzo(k)fluoranthene	1.278	1.307	-2.3	118	-0.01
37 C	Benzo(a)pyrene	1.132	1.097	3.1	113	0.00
38	Dibenzo(a,h)anthracene	1.099	1.092	0.6	114	-0.01
39	Benzo(g,h,i)perylene	1.110	1.074	3.2	112	0.00

Data Path : Z:\HPCHEM1\BNA_E\Data\BE071317\
Data File : BE093461.D
Acq On : 13 Jul 2017 16:41
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 13 17:18:19 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE070217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 03 06:40:23 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		