

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA E\DATA\BE071417\
 Data File : BE093466.D
 Acq On : 14 Jul 2017 11:39
 Operator : SJ/JU
 Sample : I4153-01
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 MW-48BR-20170711

Quant Time: Jul 17 01:52:27 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE070217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 03 06:40:23 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.93	152	4612	0.40	ng	0.00
7) Naphthalene-d8	10.72	136	19873	0.40	ng	0.00
13) Acenaphthene-d10	14.54	164	11148	0.40	ng	0.00
19) Phenanthrene-d10	17.27	188	21582	0.40	ng	0.00
27) Chrysene-d12	21.42	240	36721	0.40	ng	-0.01
34) Perylene-d12	23.93	264	31672	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.53	112	1274	0.09	ng	0.00
5) Phenol-d6	7.10	99	1190	0.06	ng	0.00
8) Nitrobenzene-d5	9.08	82	4818	0.38	ng	0.00
11) 2-Methylnaphthalene-d10	12.30	152	10186	0.32	ng	0.00
14) 2,4,6-Tribromophenol	16.03	330	481	0.14	ng	0.00
15) 2-Fluorobiphenyl	13.18	172	21188	0.49	ng	0.00
25) Fluoranthene-d10	19.28	212	135989	0.52	ng	0.00
29) Terphenyl-d14	19.87	244	29674	0.45	ng	0.00
Target Compounds						
32) Bis(2-ethylhexyl)phthalate	21.33	149	7589	0.053	ng	Qvalue # 66

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA E\DATA\BE071417\
Data File : BE093466.D
Acq On : 14 Jul 2017 11:39
Operator : SJ/JU
Sample : I4153-01
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
MW-48BR-20170711

Quant Time: Jul 17 01:52:27 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE070217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 03 06:40:23 2017
Response via : Initial Calibration

