

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA E\Data\BE071417\
 Data File : BE093463.D
 Acq On : 14 Jul 2017 9:49
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc : L00 0.2 PPM
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 17 01:35:40 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE070217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 03 06:40:23 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	141	0.00
2	1,4-Dioxane	0.847	0.752	11.2	125	0.00
3	n-Nitrosodimethylamine	0.444	0.485	-9.2	194#	0.00
4 S	2-Fluorophenol	1.207	1.135	6.0	136	0.00
5 S	Phenol-d6	1.780	1.738	2.4	147	0.00
6	bis(2-Chloroethyl)ether	1.339	1.326	1.0	143	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	151#	0.00
8 S	Nitrobenzene-d5	0.256	0.281	-9.8	180#	0.00
9	Naphthalene	4.190	4.018	4.1	149	0.00
10	Hexachlorobutadiene	0.159	0.151	5.0	146	-0.02
11 SURR	2-Methylnaphthalene-d10	0.638	0.637	0.2	156#	0.00
12	2-Methylnaphthalene	0.704	0.699	0.7	156#	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	150	0.00
14 S	2,4,6-Tribromophenol	0.126	0.081	35.7#	106	0.00
15 S	2-Fluorobiphenyl	1.553	1.520	2.1	152#	0.00
16	Acenaphthylene	1.666	1.597	4.1	152#	0.00
17	Acenaphthene	1.220	1.173	3.9	150#	0.00
18	Fluorene	1.675	1.506	10.1	140	-0.02
19 I	Phenanthrene-d10	1.000	1.000	0.0	92	0.00
20	4-Bromophenyl-phenylether	0.219	0.333	-52.1#	151#	0.00
21	Hexachlorobenzene	0.237	0.314	-32.5#	136	0.00
22	Pentachlorophenol	0.035	0.046	-31.4#	146	0.00
23	Phenanthrene	1.454	1.606	-10.5	111	0.00
24	Anthracene	0.891	0.848	4.8	93	0.00
25 SURR	Fluoranthene-d10	4.823	5.778	-19.8	120	0.00
26	Fluoranthene	1.367	1.636	-19.7	121	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	112	-0.01
28	Pvrene	1.124	1.160	-3.2	118	0.00
29 S	Terphenyl-d14	0.714	0.722	-1.1	115	0.00
30	Benzo(a)anthracene	1.080	1.055	2.3	115	0.00
31	Chrysene	1.130	1.110	1.8	112	0.00
32	Bis(2-ethylhexyl)phthalate	1.555	1.713	-10.2	151#	-0.01
33	Indeno(1,2,3-cd)pyrene	0.992	1.046	-5.4	122	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	119	0.00
35	Benzo(b)fluoranthene	1.292	1.225	5.2	115	0.00
36	Benzo(k)fluoranthene	1.278	1.253	2.0	122	0.00
37 C	Benzo(a)pyrene	1.132	1.114	1.6	123	0.00
38	Dibenzo(a,h)anthracene	1.099	1.081	1.6	121	-0.01
39	Benzo(a,h,i)perylene	1.110	1.071	3.5	120	0.00

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA E\Data\BE071417\
Data File : BE093463.D
Acq On : 14 Jul 2017 9:49
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc : L00 0.2 PPM
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 17 01:35:40 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE070217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 03 06:40:23 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			