

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA E\Data\BE071417\
 Data File : BE093463.D
 Acq On : 14 Jul 2017 9:49
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc : L00 0.2 PPM
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 17 01:35:40 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE070217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 03 06:40:23 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	141	0.00
2	1,4-Dioxane	0.400	0.355	11.3	125	0.00
3	n-Nitrosodimethylamine	0.400	0.438	-9.5	194	0.00
4 S	2-Fluorophenol	0.400	0.376	6.0	136	0.00
5 S	Phenol-d6	0.400	0.390	2.5	147	0.00
6	bis(2-Chloroethyl)ether	0.400	0.396	1.0	143	-0.01
7 I	Naphthalene-d8	0.400	0.400	0.0	151	0.00
8 S	Nitrobenzene-d5	0.400	0.440	-10.0	180	0.00
9	Naphthalene	0.400	0.384	4.0	149	0.00
10	Hexachlorobutadiene	0.400	0.379	5.3	146	-0.02
11 SURR	2-Methylnaphthalene-d10	0.400	0.399	0.3	156	0.00
12	2-Methylnaphthalene	0.400	0.397	0.8	156	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	150	0.00
14 S	2,4,6-Tribromophenol	0.400	0.257	35.8#	106	0.00
15 S	2-Fluorobiphenyl	0.400	0.392	2.0	152	0.00
16	Acenaphthylene	0.400	0.383	4.3	152	0.00
17	Acenaphthene	0.400	0.384	4.0	150	0.00
18	Fluorene	0.400	0.360	10.0	140	-0.02
19 I	Phenanthrene-d10	0.400	0.400	0.0	92	0.00
20	4-Bromophenyl-phenylether	0.400	0.608	-52.0#	151	0.00
21	Hexachlorobenzene	0.400	0.531	-32.8#	136	0.00
22	Pentachlorophenol	0.400	0.593	-48.2#	146	0.00
23	Phenanthrene	0.400	0.442	-10.5	111	0.00
24	Anthracene	0.400	0.381	4.8	93	0.00
25 SURR	Fluoranthene-d10	0.400	0.479	-19.7	120	0.00
26	Fluoranthene	0.400	0.479	-19.7	121	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	112	-0.01
28	Pvrene	0.400	0.413	-3.2	118	0.00
29 S	Terphenyl-d14	0.400	0.404	-1.0	115	0.00
30	Benzo(a)anthracene	0.400	0.391	2.3	115	0.00
31	Chrysene	0.400	0.393	1.8	112	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.441	-10.2	151	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.422	-5.5	122	-0.02
34 I	Perylene-d12	0.400	0.400	0.0	119	0.00
35	Benzo(b)fluoranthene	0.400	0.379	5.3	115	0.00
36	Benzo(k)fluoranthene	0.400	0.392	2.0	122	0.00
37 C	Benzo(a)pyrene	0.400	0.394	1.5	123	0.00
38	Dibenzo(a,h)anthracene	0.400	0.393	1.8	121	-0.01
39	Benzo(a,h,i)perylene	0.400	0.386	3.5	120	0.00

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA E\Data\BE071417\
Data File : BE093463.D
Acq On : 14 Jul 2017 9:49
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc : L00 0.2 PPM
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 17 01:35:40 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE070217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 03 06:40:23 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				