

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA E\Data\BE071417\
 Data File : BE093475.D
 Acq On : 14 Jul 2017 18:37
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 17 02:18:08 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE070217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 03 06:40:23 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	132	-0.01
2	1,4-Dioxane	0.847	0.724	14.5	113	-0.01
3	n-Nitrosodimethylamine	0.444	0.438	1.4	164#	0.00
4 S	2-Fluorophenol	1.207	1.134	6.0	127	-0.01
5 S	Phenol-d6	1.780	1.778	0.1	141	-0.01
6	bis(2-Chloroethyl)ether	1.339	1.343	-0.3	136	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	150	0.00
8 S	Nitrobenzene-d5	0.256	0.278	-8.6	176#	0.00
9	Naphthalene	4.190	4.061	3.1	149	0.00
10	Hexachlorobutadiene	0.159	0.147	7.5	141	-0.02
11 SURR	2-Methylnaphthalene-d10	0.638	0.645	-1.1	156#	0.00
12	2-Methylnaphthalene	0.704	0.706	-0.3	156#	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	150#	0.00
14 S	2,4,6-Tribromophenol	0.126	0.074	41.3#	98	0.00
15 S	2-Fluorobiphenyl	1.553	1.527	1.7	154#	0.00
16	Acenaphthylene	1.666	1.588	4.7	152#	0.00
17	Acenaphthene	1.220	1.169	4.2	150#	0.00
18	Fluorene	1.675	1.514	9.6	141	-0.02
19 I	Phenanthrene-d10	1.000	1.000	0.0	92	0.00
20	4-Bromophenyl-phenylether	0.219	0.317	-44.7#	144	0.00
21	Hexachlorobenzene	0.237	0.308	-30.0#	134	0.00
22	Pentachlorophenol	0.035	0.046	-31.4#	147	0.00
23	Phenanthrene	1.454	1.504	-3.4	104	0.00
24	Anthracene	0.891	0.829	7.0	91	0.00
25 SURR	Fluoranthene-d10	4.823	5.347	-10.9	112	0.00
26	Fluoranthene	1.367	1.527	-11.7	113	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	105	-0.01
28	Pvrene	1.124	1.160	-3.2	111	0.00
29 S	Terphenyl-d14	0.714	0.720	-0.8	108	0.00
30	Benzo(a)anthracene	1.080	0.999	7.5	102	0.00
31	Chrysene	1.130	1.111	1.7	105	0.00
32	Bis(2-ethylhexyl)phthalate	1.555	1.366	12.2	112	-0.01
33	Indeno(1,2,3-cd)pyrene	0.992	1.042	-5.0	113	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	109	-0.01
35	Benzo(b)fluoranthene	1.292	1.250	3.3	108	0.00
36	Benzo(k)fluoranthene	1.278	1.238	3.1	111	-0.01
37 C	Benzo(a)pyrene	1.132	1.092	3.5	111	0.00
38	Dibenzo(a,h)anthracene	1.099	1.094	0.5	113	-0.02
39	Benzo(a,h,i)perylene	1.110	1.079	2.8	112	-0.01

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA E\Data\BE071417\
Data File : BE093475.D
Acq On : 14 Jul 2017 18:37
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 17 02:18:08 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE070217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 03 06:40:23 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			