

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE071615\
 Data File : BE090137.D
 Acq On : 17 Jul 2015 2:20
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Jul 17 13:16:53 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE071615.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 17 12:59:41 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	79	0.00
2	1,4-Dioxane	0.667	0.663	0.6	85	0.00
3	n-Nitrosodimethylamine	0.788	0.713	9.5	75	0.00
4 S	2-Fluorophenol	1.071	1.098	-2.5	86	0.00
5 S	Phenol-d6	1.515	1.808	-19.3	99	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	85	0.00
7 S	Nitrobenzene-d5	0.370	0.436	-17.8	105	0.00
8	Nitrobenzene	0.390	0.331	15.1	75	0.00
9	Naphthalene	1.133	0.981	13.4	77	0.00
10	Hexachlorobutadiene	0.191	0.167	12.6	78	0.00
11	2-Methylnaphthalene	0.725	0.630	13.1	77	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	91	0.00
13 S	2,4,6-Tribromophenol	0.086	0.130	-51.2#	162#	0.00
14 S	2-Fluorobiphenyl	1.538	2.106	-36.9#	128	0.00
15	Acenaphthylene	8.892	7.311	17.8	78	0.00
16	Acenaphthene	1.286	1.058	17.7	78	0.00
17	Fluorene	1.680	1.353	19.5	76	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	94	0.00
19	4-Bromophenyl-phenylether	0.214	0.169	21.0	77	0.00
20	Hexachlorobenzene	0.930	0.736	20.9	75	0.00
21	Pentachlorophenol	0.036	0.024	33.3#	84	0.00
22	Phenanthrene	1.259	1.025	18.6	79	0.00
23	Anthracene	1.116	0.923	17.3	80	0.00
24	Fluoranthene	1.244	1.022	17.8	79	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	94	0.00
26	Pyrene	1.352	1.096	18.9	78	0.00
27 S	Terphenyl-d14	0.839	1.414	-68.5#	158#	0.00
28	Benzo(a)anthracene	0.848	0.731	13.8	87	0.00
29	Chrysene	1.503	1.150	23.5	78	0.00
30	Bis(2-ethylhexyl)phthalate	0.174	0.146	16.1	84	0.00
31	Indeno(1,2,3-cd)pyrene	0.265	0.279	-5.3	92	0.00
32 I	Perylene-d12	1.000	1.000	0.0	95	0.00
33	Benzo(b)fluoranthene	0.363	0.359	1.1	95	0.00
34	Benzo(k)fluoranthene	1.057	0.872	17.5	73	0.00
35 C	Benzo(a)pyrene	0.524	0.440	16.0	91	0.00
36	Dibenzo(a,h)anthracene	0.249	0.247	0.8	94	0.00
37	Benzo(g,h,i)perylene	0.552	0.497	10.0	90	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0