

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE071615\
 Data File : BE090139.D
 Acq On : 17 Jul 2015 4:07
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Jul 17 13:17:11 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE071615.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 17 12:59:41 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	76	0.00
2	1,4-Dioxane	1.000	1.110	-11.0	82	0.00
3	n-Nitrosodimethylamine	1.000	0.925	7.5	74	0.00
4 S	2-Fluorophenol	1.000	0.851	14.9	69	0.00
5 S	Phenol-d6	1.000	0.887	11.3	71	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	72	0.00
7 S	Nitrobenzene-d5	1.000	0.910	9.0	69	0.00
8	Nitrobenzene	1.000	0.912	8.8	68	0.00
9	Naphthalene	1.000	0.957	4.3	72	0.00
10	Hexachlorobutadiene	1.000	0.999	0.1	75	0.00
11	2-Methylnaphthalene	1.000	0.946	5.4	71	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	70	0.00
13 S	2,4,6-Tribromophenol	1.000	0.847	15.3	72	0.00
14 S	2-Fluorobiphenyl	1.000	0.977	2.3	70	0.00
15	Acenaphthylene	1.000	0.964	3.6	70	0.00
16	Acenaphthene	1.000	0.966	3.4	70	0.00
17	Fluorene	1.000	0.946	5.4	68	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	71	0.00
19	4-Bromophenyl-phenylether	1.000	0.957	4.3	70	0.00
20	Hexachlorobenzene	1.000	0.970	3.0	69	0.00
21	Pentachlorophenol	1.000	4.318	-331.8#	725	0.00
22	Phenanthrene	1.000	0.937	6.3	69	0.00
23	Anthracene	1.000	0.937	6.3	68	0.00
24	Fluoranthene	1.000	0.957	4.3	69	0.00
25 I	Chrysene-d12	5.000	5.000	0.0	72	0.00
26	Pyrene	1.000	0.951	4.9	70	0.00
27 S	Terphenyl-d14	1.000	0.994	0.6	71	0.00
28	Benzo(a)anthracene	1.000	1.053	-5.3	81	0.00
29	Chrysene	1.000	1.014	-1.4	71	0.00
30	Bis(2-ethylhexyl)phthalate	1.000	0.963	3.7	74	0.00
31	Indeno(1,2,3-cd)pyrene	1.000	1.308	-30.8#	103	0.00
32 I	Perylene-d12	5.000	5.000	0.0	76	0.00
33	Benzo(b)fluoranthene	1.000	1.302	-30.2#	100	0.00
34	Benzo(k)fluoranthene	1.000	0.890	11.0	74	0.00
35 C	Benzo(a)pyrene	1.000	0.940	6.0	96	0.00
36	Dibenzo(a,h)anthracene	1.000	1.272	-27.2#	106	0.00
37	Benzo(g,h,i)perylene	1.000	0.974	2.6	96	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0