

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE071715\
 Data File : BE090159.D
 Acq On : 17 Jul 2015 23:24
 Operator : TP/IZ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE071715

Quant Time: Jul 18 00:45:10 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE071715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jul 18 00:40:53 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	78	0.00
2	1,4-Dioxane	0.619	0.622	-0.5	81	0.00
3	n-Nitrosodimethylamine	0.886	0.892	-0.7	81	0.00
4 S	2-Fluorophenol	1.176	1.149	2.3	77	0.00
5 S	Phenol-d6	1.516	1.466	3.3	75	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	77	0.00
7 S	Nitrobenzene-d5	0.332	0.319	3.9	76	0.00
8	Nitrobenzene	0.327	0.326	0.3	76	0.00
9	Naphthalene	1.072	1.053	1.8	78	0.00
10	Hexachlorobutadiene	0.132	0.134	-1.5	80	0.00
11	2-Methylnaphthalene	0.720	0.699	2.9	77	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	76	0.00
13 S	2,4,6-Tribromophenol	0.118	0.101	14.4	67	0.00
14 S	2-Fluorobiphenyl	1.472	1.457	1.0	76	0.00
15	Acenaphthylene	9.006	8.857	1.7	76	0.00
16	Acenaphthene	1.267	1.230	2.9	75	0.00
17	Fluorene	1.713	1.635	4.6	74	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	74	0.00
19	4-Bromophenyl-phenylether	0.186	0.182	2.2	75	0.00
20	Hexachlorobenzene	0.659	0.642	2.6	74	0.00
21	Pentachlorophenol	0.059	0.044	25.4#	59	0.00
22	Phenanthrene	1.121	1.108	1.2	75	0.00
23	Anthracene	1.042	1.006	3.5	74	0.00
24	Fluoranthene	1.258	1.226	2.5	74	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	75	0.00
26	Pyrene	1.133	1.081	4.6	73	0.00
27 S	Terphenyl-d14	0.837	0.800	4.4	73	0.00
28	Benzo(a)anthracene	1.144	1.061	7.3	73	0.00
29	Chrysene	1.155	1.122	2.9	75	0.00
30	Bis(2-ethylhexyl)phthalate	0.587	0.480	18.2	63	0.00
31	Indeno(1,2,3-cd)pyrene	1.352	1.257	7.0	71	0.00
32 I	Perylene-d12	1.000	1.000	0.0	74	0.00
33	Benzo(b)fluoranthene	1.044	0.982	5.9	71	0.00
34	Benzo(k)fluoranthene	1.192	1.167	2.1	74	0.00
35 C	Benzo(a)pyrene	1.011	0.958	5.2	71	0.00
36	Dibenzo(a,h)anthracene	1.066	0.991	7.0	71	0.00
37	Benzo(g,h,i)perylene	1.102	1.055	4.3	72	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0