

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE071715\
 Data File : BE090159.D
 Acq On : 17 Jul 2015 23:24
 Operator : TP/IZ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE071715

Quant Time: Jul 18 00:45:10 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE071715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jul 18 00:40:53 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	78	0.00
2	1,4-Dioxane	1.000	1.120	-12.0	81	0.00
3	n-Nitrosodimethylamine	1.000	1.007	-0.7	81	0.00
4 S	2-Fluorophenol	1.000	0.978	2.2	77	0.00
5 S	Phenol-d6	1.000	0.967	3.3	75	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	77	0.00
7 S	Nitrobenzene-d5	1.000	0.961	3.9	76	0.00
8	Nitrobenzene	1.000	0.999	0.1	76	0.00
9	Naphthalene	1.000	0.983	1.7	78	0.00
10	Hexachlorobutadiene	1.000	1.013	-1.3	80	0.00
11	2-Methylnaphthalene	1.000	0.971	2.9	77	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	76	0.00
13 S	2,4,6-Tribromophenol	1.000	0.854	14.6	67	0.00
14 S	2-Fluorobiphenyl	1.000	0.990	1.0	76	0.00
15	Acenaphthylene	1.000	0.983	1.7	76	0.00
16	Acenaphthene	1.000	0.971	2.9	75	0.00
17	Fluorene	1.000	0.955	4.5	74	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	74	0.00
19	4-Bromophenyl-phenylether	1.000	0.981	1.9	75	0.00
20	Hexachlorobenzene	1.000	0.974	2.6	74	0.00
21	Pentachlorophenol	1.000	0.799	20.1	59	0.00
22	Phenanthrene	1.000	0.988	1.2	75	0.00
23	Anthracene	1.000	0.965	3.5	74	0.00
24	Fluoranthene	1.000	0.975	2.5	74	0.00
25 I	Chrysene-d12	5.000	5.000	0.0	75	0.00
26	Pyrene	1.000	0.954	4.6	73	0.00
27 S	Terphenyl-d14	1.000	0.956	4.4	73	0.00
28	Benzo(a)anthracene	1.000	0.927	7.3	73	0.00
29	Chrysene	1.000	0.971	2.9	75	0.00
30	Bis(2-ethylhexyl)phthalate	1.000	0.818	18.2	63	0.00
31	Indeno(1,2,3-cd)pyrene	1.000	0.929	7.1	71	0.00
32 I	Perylene-d12	5.000	5.000	0.0	74	0.00
33	Benzo(b)fluoranthene	1.000	0.941	5.9	71	0.00
34	Benzo(k)fluoranthene	1.000	0.979	2.1	74	0.00
35 C	Benzo(a)pyrene	1.000	0.948	5.2	71	0.00
36	Dibenzo(a,h)anthracene	1.000	0.930	7.0	71	0.00
37	Benzo(g,h,i)perylene	1.000	0.957	4.3	72	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0