

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE071715\
 Data File : BE090165.D
 Acq On : 18 Jul 2015 2:59
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Jul 18 04:52:29 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE071715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jul 18 00:40:53 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	82	0.00
2	1,4-Dioxane	1.000	1.116	-11.6	85	0.00
3	n-Nitrosodimethylamine	1.000	1.002	-0.2	84	0.00
4 S	2-Fluorophenol	1.000	0.911	8.9	75	0.00
5 S	Phenol-d6	1.000	0.940	6.0	76	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	80	0.00
7 S	Nitrobenzene-d5	1.000	0.933	6.7	77	0.00
8	Nitrobenzene	1.000	0.959	4.1	76	0.00
9	Naphthalene	1.000	0.983	1.7	81	0.00
10	Hexachlorobutadiene	1.000	0.970	3.0	79	0.00
11	2-Methylnaphthalene	1.000	0.951	4.9	78	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	76	0.00
13 S	2,4,6-Tribromophenol	1.000	0.752	24.8	59	0.00
14 S	2-Fluorobiphenyl	1.000	0.988	1.2	76	0.00
15	Acenaphthylene	1.000	0.973	2.7	76	0.00
16	Acenaphthene	1.000	0.970	3.0	76	0.00
17	Fluorene	1.000	0.950	5.0	74	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	76	0.00
19	4-Bromophenyl-phenylether	1.000	0.945	5.5	75	0.00
20	Hexachlorobenzene	1.000	0.945	5.5	74	0.00
21	Pentachlorophenol	1.000	0.751	24.9	55	0.00
22	Phenanthrene	1.000	0.968	3.2	76	0.00
23	Anthracene	1.000	0.941	5.9	74	0.02
24	Fluoranthene	1.000	0.955	4.5	75	0.00
25 I	Chrysene-d12	5.000	5.000	0.0	77	0.00
26	Pyrene	1.000	0.956	4.4	75	0.00
27 S	Terphenyl-d14	1.000	0.952	4.8	75	0.00
28	Benzo(a)anthracene	1.000	0.921	7.9	74	0.00
29	Chrysene	1.000	0.981	1.9	78	0.00
30	Bis(2-ethylhexyl)phthalate	1.000	0.713	28.7#	56	0.00
31	Indeno(1,2,3-cd)pyrene	1.000	0.889	11.1	70	0.00
32 I	Perylene-d12	5.000	5.000	0.0	76	0.00
33	Benzo(b)fluoranthene	1.000	0.927	7.3	72	0.00
34	Benzo(k)fluoranthene	1.000	0.959	4.1	74	0.00
35 C	Benzo(a)pyrene	1.000	0.905	9.5	70	0.00
36	Dibenzo(a,h)anthracene	1.000	0.887	11.3	70	0.00
37	Benzo(g,h,i)perylene	1.000	0.919	8.1	71	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0