

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE071715\
 Data File : BE090173.D
 Acq On : 18 Jul 2015 9:00
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Jul 20 04:58:58 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE071715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jul 18 00:40:53 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	93	0.00
2	1,4-Dioxane	0.619	0.611	1.3	95	0.00
3	n-Nitrosodimethylamine	0.886	0.886	0.0	95	0.00
4 S	2-Fluorophenol	1.176	1.112	5.4	89	0.00
5 S	Phenol-d6	1.516	1.433	5.5	87	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	91	0.00
7 S	Nitrobenzene-d5	0.332	0.310	6.6	87	0.00
8	Nitrobenzene	0.327	0.314	4.0	87	0.00
9	Naphthalene	1.072	1.039	3.1	91	0.00
10	Hexachlorobutadiene	0.132	0.127	3.8	89	0.00
11	2-Methylnaphthalene	0.720	0.686	4.7	89	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	88	0.00
13 S	2,4,6-Tribromophenol	0.118	0.097	17.8	74	0.00
14 S	2-Fluorobiphenyl	1.472	1.448	1.6	88	0.00
15	Acenaphthylene	9.006	8.729	3.1	88	0.00
16	Acenaphthene	1.267	1.223	3.5	88	0.00
17	Fluorene	1.713	1.591	7.1	84	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	89	0.00
19	4-Bromophenyl-phenylether	0.186	0.176	5.4	87	0.00
20	Hexachlorobenzene	0.659	0.612	7.1	85	0.00
21	Pentachlorophenol	0.059	0.049	16.9	78	0.00
22	Phenanthrene	1.121	1.102	1.7	89	0.00
23	Anthracene	1.042	0.999	4.1	88	0.00
24	Fluoranthene	1.258	1.201	4.5	87	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	87	0.00
26	Pyrene	1.133	1.104	2.6	87	0.00
27 S	Terphenyl-d14	0.837	0.803	4.1	85	0.00
28	Benzo(a)anthracene	1.144	1.060	7.3	85	0.00
29	Chrysene	1.155	1.116	3.4	87	0.00
30	Bis(2-ethylhexyl)phthalate	0.587	0.433	26.2#	66	0.00
31	Indeno(1,2,3-cd)pyrene	1.352	1.165	13.8	77	0.00
32 I	Perylene-d12	1.000	1.000	0.0	84	0.00
33	Benzo(b)fluoranthene	1.044	0.974	6.7	80	0.00
34	Benzo(k)fluoranthene	1.192	1.146	3.9	83	0.00
35 C	Benzo(a)pyrene	1.011	0.922	8.8	78	0.00
36	Dibenzo(a,h)anthracene	1.066	0.942	11.6	77	0.00
37	Benzo(g,h,i)perylene	1.102	0.999	9.3	78	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0