

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE071715\
 Data File : BE090158.D
 Acq On : 17 Jul 2015 22:48
 Operator : TP/IZ
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.1

Manual Integrations
 APPROVED

apatel
 7/20/2015 3:39:26 PM

Quant Time: Jul 18 00:26:07 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE071715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 17 23:58:28 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.61	152	10031	5.00	ng	0.00
6) Naphthalene-d8	10.36	136	46664	5.00	ng	0.00
12) Acenaphthene-d10	14.22	164	26429	5.00	ng	0.00
18) Phenanthrene-d10	16.96	188	67013	5.00	ng	0.00
25) Chrysene-d12	21.11	240	72777	5.00	ng	0.00
32) Perylene-d12	23.43	264	74824	5.00	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.24	112	270	0.12	ng	0.00
5) Phenol-d6	6.80	99	303	0.10	ng	0.00
7) Nitrobenzene-d5	8.76	82	339	0.10	ng	0.00
13) 2,4,6-Tribromophenol	15.73	330	56	0.12	ng	0.00
14) 2-Fluorobiphenyl	12.85	172	946	0.12	ng	0.00
27) Terphenyl-d14	19.58	244	1448	0.12	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.22	88	174	0.13	ng	90
3) n-Nitrosodimethylamine	3.52	42	204	0.13	ng	# 89
8) Nitrobenzene	8.79	77	308	0.09	ng	99
9) Naphthalene	10.41	128	1264	0.12	ng	# 90
10) Hexachlorobutadiene	10.71	225	156	0.09	ng	98
11) 2-Methylnaphthalene	12.04	142	812	0.12	ng	# 94
15) Acenaphthylene	13.94	152	5777	0.12	ng	100
16) Acenaphthene	14.28	154	818	0.12	ng	96
17) Fluorene	15.27	166	1135	0.13	ng	95
19) 4-Bromophenyl-phenylether	16.16	248	307	0.11	ng	97
20) Hexachlorobenzene	16.28	284	1173	0.10	ng	93
21) Pentachlorophenol	16.63	266	54	0.10	ng	# 73
22) Phenanthrene	17.00	178	1872	0.11	ng	97
23) Anthracene	17.10	178	1587	0.11	ng	99
24) Fluoranthene	19.00	202	1895	0.11	ng	98
26) Pyrene	19.37	202	2030	0.11	ng	100
28) Benzo(a)anthracene	21.10	228	2101	0.16	ng	97
29) Chrysene	21.15	228	2034	0.10	ng	95
30) Bis(2-ethylhexyl)phthalate	21.05	149	877	0.25	ng	99
31) Indeno(1,2,3-cd)pyrene	25.83	276	2185	0.34	ng	100
33) Benzo(b)fluoranthene	22.73	252	1733	0.24	ng	# 89
34) Benzo(k)fluoranthene	22.77	252	2159m	0.14	ng	
35) Benzo(a)pyrene	23.32	252	1709	0.19	ng	# 86
36) Dibenzo(a,h)anthracene	25.85	278	1714	0.30	ng	# 81
37) Benzo(g,h,i)perylene	26.56	276	1936	0.21	ng	# 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA_E\DATA\BE071715\
Data File : BE090158.D
Acq On : 17 Jul 2015 22:48
Operator : TP/IZ
Sample : SSTDICC0.1
Misc :
ALS Vial : 9 Sample Multiplier: 1


Instrument :
BNA_E
ClientSampleId :
SSTDICC0.1

Manual Integrations
APPROVED

apatel
7/20/2015 3:39:26 PM

Quant Time: Jul 18 00:26:07 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE071715.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 17 23:58:28 2015
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 7.61 min Scan# 850

Delta R.T. 0.00 min

Lab File: BE090158.D

Acq: 17 Jul 2015 22:48

Instrument :

BNA_E





ClientSampleId :

SSTDICC0.1

Tgt Ion:	152	Resp:	10031
Ion Ratio	Lower	Upper	
152	100		
150	157.3	127.2	190.8
115	64.4	51.6	77.4

Manual Integrations
APPROVED

apatel
7/20/2015 3:39:26 PM



#2

1,4-Dioxane

Concen: 0.13 ng




RT: 3.22 min Scan# 92


Delta R.T. 0.01 min

Lab File: BE090158.D

Acq: 17 Jul 2015 22:48

Tgt Ion:	88	Resp:	174
Ion Ratio	Lower	Upper	
88	100		
58	63.2	58.2	87.2
43	38.5	27.7	41.5





#3





n-Nitrosodimethylamine
Concen: 0.13 ng
RT: 3.52 min Scan# 135
Delta R.T. 0.01 min
Lab File: BE090158.D
Acq: 17 Jul 2015 22:48

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1

Tgt Ion	Ratio	Lower	Upper
42	100		
74	95.6	82.2	123.4
44	29.4	7.5	11.3#

Manual Integrations
APPROVED




apatel
7/20/2015 3:39:26 PM




#4

2-Fluorophenol
Concen: 0.12 ng
RT: 5.24 min Scan# 389
Delta R.T. -0.00 min
Lab File: BE090158.D
Acq: 17 Jul 2015 22:48

Tgt Ion	Ratio	Lower	Upper
112	100		
64	59.3	46.7	70.1
63	38.1	27.3	40.9





#5

Phenol-d6

Concen: 0.10 ng

RT: 6.80 min Scan# 672

Delta R.T. 0.01 min

Lab File: BE090158.D

Acq: 17 Jul 2015 22:48

Instrument :

BNA_E





ClientSampleId :

SSTDIC0.1

Tgt Ion:	99	Resp:	303
Ion	Ratio	Lower	Upper
99	100		
42	20.5	19.1	28.7
71	37.6	29.4	44.2

Manual Integrations
APPROVED

apatel
7/20/2015 3:39:26 PM



#6

Naphthalene-d8

Concen: 5.00 ng




RT: 10.36 min Scan# 1374


Delta R.T. -0.00 min

Lab File: BE090158.D

Acq: 17 Jul 2015 22:48

Tgt Ion:	136	Resp:	46664
Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.9	13.3
54	8.6	6.3	9.5
68	5.9	4.3	6.5





#7

Nitrobenzene-d5

Concen: 0.10 ng

RT: 8.76 min Scan# 1099

Delta R.T. 0.01 min

Lab File: BE090158.D

Acq: 17 Jul 2015 22:48

Instrument :

BNA_E





ClientSampleId :

SSTDICC0.1

Tgt Ion:	82	Resp:	339
Ion	Ratio	Lower	Upper
82	100		
128	55.2	32.7	49.1#
54	66.3	45.9	68.9

Manual Integrations
APPROVED

apatel
7/20/2015 3:39:26 PM



#8

Nitrobenzene

Concen: 0.09 ng




RT: 8.79 min Scan# 1107


Delta R.T. 0.00 min

Lab File: BE090158.D

Acq: 17 Jul 2015 22:48

Tgt Ion:	77	Resp:	308
Ion	Ratio	Lower	Upper
77	100		
123	40.6	32.3	48.5
65	13.6	10.4	15.6





#9

Naphthalene

Concen: 0.12 ng

RT: 10.41 min Scan# 1381

Delta R.T. -0.00 min

Lab File: BE090158.D

Acq: 17 Jul 2015 22:48

Instrument :

BNA_E


ClientSampleId :

SSTDIC0.1




Tgt Ion:	128	Resp:	1264
Ion Ratio	Lower	Upper	
128	100		
129	16.3	9.8	14.6#
127	17.5	10.7	16.1#

Manual Integrations
APPROVED

apatel
7/20/2015 3:39:26 PM



Abundance Ion 128.00 (127.70 to 128.70): E
Ion 129.00 (128.70 to 129.70): E
Ion 127.00 (126.70 to 127.70): E



#10

Hexachlorobutadiene

Concen: 0.09 ng


RT: 10.71 min Scan# 1420

Delta R.T. -0.00 min



Lab File: BE090158.D


Acq: 17 Jul 2015 22:48

Tgt Ion:	225	Resp:	156
Ion Ratio	Lower	Upper	
225	100		
223	64.7	50.3	75.5
227	62.2	51.3	76.9



Abundance Ion 225.00 (224.70 to 225.70): E
Ion 223.00 (222.70 to 223.70): E
Ion 227.00 (226.70 to 227.70): E









#11
2-Methylnaphthalene
Concen: 0.12 ng
RT: 12.04 min Scan# 1656
Delta R.T. 0.01 min
Lab File: BE090158.D
Acq: 17 Jul 2015 22:48

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1

Tgt Ion	Ratio	Lower	Upper
142	100		
141	95.0	73.9	110.9
115	46.8	30.0	45.0




Manual Integrations
APPROVED


apatel
7/20/2015 3:39:26 PM



#12
Acenaphthene-d10
Concen: 5.00 ng
RT: 14.22 min Scan# 2078
Delta R.T. -0.00 min
Lab File: BE090158.D
Acq: 17 Jul 2015 22:48

Tgt Ion	Ratio	Lower	Upper
164	100		
162	107.3	83.3	124.9
160	54.3	40.9	61.3









#13
 2,4,6-Tribromophenol
 Concen: 0.12 ng
 RT: 15.73 min Scan# 2207
 Delta R.T. -0.00 min
 Lab File: BE090158.D
 Acq: 17 Jul 2015 22:48

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.1

Tgt Ion	Ratio	Lower	Upper
330	100		
332	121.4	78.2	117.2#
141	353.6	215.8	323.6#




Manual Integrations
 APPROVED


apatel
 7/20/2015 3:39:26 PM



#14
 2-Fluorobiphenyl
 Concen: 0.12 ng
 RT: 12.85 min Scan# 1834
 Delta R.T. 0.01 min
 Lab File: BE090158.D
 Acq: 17 Jul 2015 22:48

Tgt Ion	Ratio	Lower	Upper
172	100		
171	39.8	28.5	42.7
170	30.7	19.7	29.5#





#15

Acenaphthylene

Concen: 0.12 ng

RT: 13.94 min Scan# 2035

Delta R.T. 0.01 min

Lab File: BE090158.D

Acq: 17 Jul 2015 22:48

Instrument :

BNA_E





ClientSampleId :

SSTDIC0.1

Tgt Ion:	152	Resp:	5777
Ion Ratio	Lower	Upper	
152	100		
151	5.0	4.0	6.0
153	12.9	10.4	15.6

Manual Integrations
APPROVED

apatel
7/20/2015 3:39:26 PM



#16

Acenaphthene

Concen: 0.12 ng




RT: 14.28 min Scan# 2088


Delta R.T. 0.01 min

Lab File: BE090158.D

Acq: 17 Jul 2015 22:48

Tgt Ion:	154	Resp:	818
Ion Ratio	Lower	Upper	
154	100		
153	475.4	376.9	565.3
152	261.4	199.0	298.6





#17

Fluorene

Concen: 0.13 ng

RT: 15.27 min Scan# 2181

Delta R.T. -0.00 min

Lab File: BE090158.D

Acq: 17 Jul 2015 22:48

Instrument :

BNA_E





ClientSampleId :

SSTDICC0.1

Tgt Ion:	166	Resp:	1135
Ion Ratio	Lower	Upper	
166	100		
165	96.3	81.5	122.3
167	14.6	11.9	17.9

Manual Integrations
APPROVED

apatel
7/20/2015 3:39:26 PM



#18

Phenanthrene-d10

Concen: 5.00 ng




RT: 16.96 min Scan# 2278


Delta R.T. -0.00 min

Lab File: BE090158.D

Acq: 17 Jul 2015 22:48

Tgt Ion:	188	Resp:	67013
Ion Ratio	Lower	Upper	
188	100		
94	6.1	4.6	7.0
80	6.3	4.6	6.8









#19
4-Bromophenyl-phenylether
Concen: 0.11 ng
RT: 16.16 min Scan# 2232
Delta R.T. -0.00 min
Lab File: BE090158.D
Acq: 17 Jul 2015 22:48

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1

Tgt Ion	Ratio	Lower	Upper
248	100		
250	94.5	77.4	116.0
141	358.3	281.3	421.9




Manual Integrations
APPROVED


apatel
7/20/2015 3:39:26 PM



#20
Hexachlorobenzene
Concen: 0.10 ng
RT: 16.28 min Scan# 2239
Delta R.T. -0.00 min
Lab File: BE090158.D
Acq: 17 Jul 2015 22:48

Tgt Ion	Ratio	Lower	Upper
284	100		
142	59.4	42.2	63.2
249	29.6	25.5	38.3









#21
 Pentachlorophenol
 Concen: 0.10 ng
 RT: 16.63 min Scan# 2259
 Delta R.T. -0.00 min
 Lab File: BE090158.D
 Acq: 17 Jul 2015 22:48

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.1

Tgt Ion	Ratio	Lower	Upper
266	100		
268	96.6	58.2	87.4#
264	89.8	55.0	82.4#




Manual Integrations
 APPROVED


apatel
 7/20/2015 3:39:26 PM



#22
 Phenanthrene
 Concen: 0.11 ng
 RT: 17.00 min Scan# 2280
 Delta R.T. -0.00 min
 Lab File: BE090158.D
 Acq: 17 Jul 2015 22:48

Tgt Ion	Ratio	Lower	Upper
178	100		
176	20.5	15.9	23.9
179	18.1	12.5	18.7





#23

Anthracene

Concen: 0.11 ng

RT: 17.10 min Scan# 2286

Delta R.T. 0.02 min

Lab File: BE090158.D

Acq: 17 Jul 2015 22:48

Instrument :

BNA_E


ClientSampled :

SSTDIC0.1




Tgt Ion:	178	Resp:	1587
Ion Ratio	Lower	Upper	
178	100		
176	19.4	15.4	23.2
179	15.9	12.2	18.2

Manual Integrations
APPROVED

apatel
7/20/2015 3:39:26 PM



Abundance Ion 178.00 (177.70 to 178.70): E
Ion 176.00 (175.70 to 176.70): E
Ion 179.00 (178.70 to 179.70): E



#24

Fluoranthene

Concen: 0.11 ng

RT: 19.00 min Scan# 2509




Delta R.T. -0.00 min


Lab File: BE090158.D

Acq: 17 Jul 2015 22:48

Tgt Ion:	202	Resp:	1895
Ion Ratio	Lower	Upper	
202	100		
101	11.0	8.5	12.7
203	18.2	13.9	20.9

Abundance Ion 202.00 (201.70 to 202.70): E
Ion 101.00 (100.70 to 101.70): E
Ion 203.00 (202.70 to 203.70): E





#25

Chrysene-d12

Concen: 5.00 ng

RT: 21.11 min Scan# 2867

Delta R.T. 0.00 min

Lab File: BE090158.D


Acq: 17 Jul 2015 22:48

Instrument :

BNA_E

ClientSampleId :




SSTDIC0.1



Tgt Ion:	240	Resp:	72777
Ion Ratio	Lower	Upper	
240	100		
120	11.1	9.4	14.0
236	27.1	21.8	32.8

Manual Integrations
APPROVED

apatel
7/20/2015 3:39:26 PM



#26

Pyrene


Concen: 0.11 ng

RT: 19.37 min Scan# 2578



Delta R.T. 0.00 min


Lab File: BE090158.D

Acq: 17 Jul 2015 22:48



Tgt Ion:	202	Resp:	2030
Ion Ratio	Lower	Upper	
202	100		
200	21.2	17.0	25.6
203	17.7	14.1	21.1









#27
 Terphenyl-d14
 Concen: 0.12 ng
 RT: 19.58 min Scan# 2618
 Delta R.T. 0.00 min
 Lab File: BE090158.D
 Acq: 17 Jul 2015 22:48

Instrument :
 BNA_E
 ClientSampled :
 SSTDICC0.1

Tgt Ion	Ratio	Lower	Upper
244	100		
212	11.5	6.5	9.7#
122	14.0	9.8	14.6




Manual Integrations
 APPROVED


apatel
 7/20/2015 3:39:26 PM



#28
 Benzo(a)anthracene
 Concen: 0.16 ng
 RT: 21.10 min Scan# 2865
 Delta R.T. 0.00 min
 Lab File: BE090158.D
 Acq: 17 Jul 2015 22:48

Tgt Ion	Ratio	Lower	Upper
228	100		
226	28.5	22.2	33.4
229	22.1	15.7	23.5









#29
Chrysene
Concen: 0.10 ng
RT: 21.15 min Scan# 2872
Delta R.T. -0.01 min
Lab File: BE090158.D
Acq: 17 Jul 2015 22:48

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1

Tgt Ion:	228	Resp:	2034
Ion Ratio	Lower	Upper	
228	100		
226	33.1	23.8	35.8
229	21.8	16.5	24.7




Manual Integrations
APPROVED


apatel
7/20/2015 3:39:26 PM



#30
Bis(2-ethylhexyl)phthalate
Concen: 0.25 ng
RT: 21.05 min Scan# 2857
Delta R.T. 0.00 min
Lab File: BE090158.D
Acq: 17 Jul 2015 22:48

Tgt Ion:	149	Resp:	877
Ion Ratio	Lower	Upper	
149	100		
167	23.0	18.6	28.0
279	2.7	2.2	3.4









#31
 Indeno(1,2,3-cd)pyrene
 Concen: 0.34 ng
 RT: 25.83 min Scan# 3837
 Delta R.T. -0.00 min
 Lab File: BE090158.D
 Acq: 17 Jul 2015 22:48

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.1

Tgt Ion	Ratio	Resp	Lower	Upper
276	100	2185		
138	24.2	19.5	29.3	
277	24.4	19.8	29.6	




Manual Integrations
 APPROVED


apatel
 7/20/2015 3:39:26 PM



#32
 Perylene-d12
 Concen: 5.00 ng
 RT: 23.43 min Scan# 3314
 Delta R.T. -0.00 min
 Lab File: BE090158.D
 Acq: 17 Jul 2015 22:48

Tgt Ion	Ratio	Resp	Lower	Upper
264	100	74824		
260	24.6	19.8	29.8	
265	21.6	17.4	26.2	









#33
Benzo(b)fluoranthene
Concen: 0.24 ng
RT: 22.73 min Scan# 3160
Delta R.T. 0.00 min
Lab File: BE090158.D
Acq: 17 Jul 2015 22:48

Instrument :
BNA_E
ClientSampleId :
SSTDICC0.1

Tgt Ion	Ratio	Lower	Upper
252	100		
253	26.7	17.9	26.9
125	16.9	8.7	13.1#




Manual Integrations
APPROVED


apatel
7/20/2015 3:39:26 PM



#34
Benzo(k)fluoranthene
Concen: 0.14 ng m
RT: 22.77 min Scan# 3170
Delta R.T. 0.00 min
Lab File: BE090158.D
Acq: 17 Jul 2015 22:48

Tgt Ion	Ratio	Lower	Upper
252	100		
253	27.1	18.1	27.1
125	17.8	8.7	13.1#









#35
Benzo(a)pyrene
Concen: 0.19 ng
RT: 23.32 min Scan# 3291
Delta R.T. -0.00 min
Lab File: BE090158.D
Acq: 17 Jul 2015 22:48

Instrument :
BNA_E
ClientSampleId :
SSTDICC0.1

Tgt Ion	Ratio	Lower	Upper
252	100		
253	28.0	18.0	27.0#
125	18.7	9.4	14.0#




Manual Integrations
APPROVED


apatel
7/20/2015 3:39:26 PM



#36
Dibenzo(a,h)anthracene
Concen: 0.30 ng
RT: 25.85 min Scan# 3842
Delta R.T. 0.00 min
Lab File: BE090158.D
Acq: 17 Jul 2015 22:48

Tgt Ion	Ratio	Lower	Upper
278	100		
139	26.0	12.9	19.3#
279	31.9	19.3	28.9#





#37

Benzo(g,h,i)perylene

Concen: 0.21 ng

RT: 26.56 min Scan# 3998

Delta R.T. -0.00 min

Lab File: BE090158.D


Acq: 17 Jul 2015 22:48

Instrument :

BNA_E

ClientSampleId :

SSTDIC0.1



Tgt Ion: 276 Resp: 1936

Ion Ratio Lower Upper


276 100

277 28.4 19.0 28.6

138 28.1 17.2 25.8#

Manual Integrations
APPROVED

apatel
7/20/2015 3:39:26 PM



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

