

Data Path : S:\HPCHEM1\BNA_E\DATA\BE072015\
 Data File : BE090175.D
 Acq On : 20 Jul 2015 14:04
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Jul 21 02:09:46 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE071715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jul 18 00:40:53 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	125	0.00
2	1,4-Dioxane	0.619	0.620	-0.2	130	0.00
3	n-Nitrosodimethylamine	0.886	0.879	0.8	128	0.00
4 S	2-Fluorophenol	1.176	1.066	9.4	115	-0.01
5 S	Phenol-d6	1.516	1.424	6.1	117	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	121	0.00
7 S	Nitrobenzene-d5	0.332	0.314	5.4	118	0.00
8	Nitrobenzene	0.327	0.320	2.1	118	0.00
9	Naphthalene	1.072	1.038	3.2	121	0.00
10	Hexachlorobutadiene	0.132	0.128	3.0	120	0.00
11	2-Methylnaphthalene	0.720	0.688	4.4	119	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	118	0.00
13 S	2,4,6-Tribromophenol	0.118	0.079	33.1#	81	-0.02
14 S	2-Fluorobiphenyl	1.472	1.437	2.4	117	0.00
15	Acenaphthylene	9.006	8.540	5.2	115	0.00
16	Acenaphthene	1.267	1.196	5.6	114	0.00
17	Fluorene	1.713	1.598	6.7	113	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	113	0.00
19	4-Bromophenyl-phenylether	0.186	0.184	1.1	116	0.00
20	Hexachlorobenzene	0.659	0.615	6.7	108	0.00
21	Pentachlorophenol	0.059	0.044	25.4#	90	0.00
22	Phenanthrene	1.121	1.103	1.6	114	0.00
23	Anthracene	1.042	1.013	2.8	113	0.00
24	Fluoranthene	1.258	1.215	3.4	112	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	115	0.00
26	Pyrene	1.133	1.078	4.9	112	0.00
27 S	Terphenyl-d14	0.837	0.804	3.9	112	0.00
28	Benzo(a)anthracene	1.144	1.051	8.1	111	0.00
29	Chrysene	1.155	1.095	5.2	112	0.00
30	Bis(2-ethylhexyl)phthalate	0.587	0.355	39.5#	71	0.00
31	Indeno(1,2,3-cd)pyrene	1.352	1.104	18.3	96	0.00
32 I	Perylene-d12	1.000	1.000	0.0	109	0.00
33	Benzo(b)fluoranthene	1.044	0.952	8.8	102	0.00
34	Benzo(k)fluoranthene	1.192	1.136	4.7	106	0.00
35 C	Benzo(a)pyrene	1.011	0.905	10.5	100	0.00
36	Dibenzo(a,h)anthracene	1.066	0.902	15.4	95	0.00
37	Benzo(g,h,i)perylene	1.102	0.960	12.9	97	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0