

Data Path : S:\HPCHEM1\BNA_E\DATA\BE072015\
 Data File : BE090175.D
 Acq On : 20 Jul 2015 14:04
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Jul 21 02:09:46 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE071715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jul 18 00:40:53 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	125	0.00
2	1,4-Dioxane	1.000	1.116	-11.6	130	0.00
3	n-Nitrosodimethylamine	1.000	0.992	0.8	128	0.00
4 S	2-Fluorophenol	1.000	0.907	9.3	115	-0.01
5 S	Phenol-d6	1.000	0.939	6.1	117	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	121	0.00
7 S	Nitrobenzene-d5	1.000	0.947	5.3	118	0.00
8	Nitrobenzene	1.000	0.980	2.0	118	0.00
9	Naphthalene	1.000	0.969	3.1	121	0.00
10	Hexachlorobutadiene	1.000	0.972	2.8	120	0.00
11	2-Methylnaphthalene	1.000	0.955	4.5	119	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	118	0.00
13 S	2,4,6-Tribromophenol	1.000	0.671	32.9#	81	-0.02
14 S	2-Fluorobiphenyl	1.000	0.976	2.4	117	0.00
15	Acenaphthylene	1.000	0.948	5.2	115	0.00
16	Acenaphthene	1.000	0.944	5.6	114	0.00
17	Fluorene	1.000	0.932	6.8	113	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	113	0.00
19	4-Bromophenyl-phenylether	1.000	0.991	0.9	116	0.00
20	Hexachlorobenzene	1.000	0.933	6.7	108	0.00
21	Pentachlorophenol	1.000	0.800	20.0	90	0.00
22	Phenanthrene	1.000	0.984	1.6	114	0.00
23	Anthracene	1.000	0.972	2.8	113	0.00
24	Fluoranthene	1.000	0.966	3.4	112	0.00
25 I	Chrysene-d12	5.000	5.000	0.0	115	0.00
26	Pyrene	1.000	0.952	4.8	112	0.00
27 S	Terphenyl-d14	1.000	0.960	4.0	112	0.00
28	Benzo(a)anthracene	1.000	0.919	8.1	111	0.00
29	Chrysene	1.000	0.948	5.2	112	0.00
30	Bis(2-ethylhexyl)phthalate	1.000	0.604	39.6#	71	0.00
31	Indeno(1,2,3-cd)pyrene	1.000	0.817	18.3	96	0.00
32 I	Perylene-d12	5.000	5.000	0.0	109	0.00
33	Benzo(b)fluoranthene	1.000	0.912	8.8	102	0.00
34	Benzo(k)fluoranthene	1.000	0.953	4.7	106	0.00
35 C	Benzo(a)pyrene	1.000	0.895	10.5	100	0.00
36	Dibenzo(a,h)anthracene	1.000	0.847	15.3	95	0.00
37	Benzo(g,h,i)perylene	1.000	0.871	12.9	97	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0