

#17

Acenaphthene

Concen: 0.233 ng

RT: 14.58 min Scan# 1116

Delta R.T. -0.00 min


Lab File: BE093506.D

Acq: 21 Jul 2017 13:07

Instrument :

BNA_E

ClientSampleId :






Tgt Ion:154 Resp: 5950

Ion Ratio Lower Upper

154 100

153 114.5 91.2 136.8

152 55.6 44.8 67.2



#18

Fluorene

Concen: 0.240 ng

RT: 15.58 min Scan# 1183

Delta R.T. 0.01 min

Lab File: BE093506.D

Acq: 21 Jul 2017 13:07

Tgt Ion:166 Resp: 8412

Ion Ratio Lower Upper

166 100

165 98.0 78.6 117.8

167 55.0 11.1 16.7#

