

#32

Bis(2-ethylhexyl)phthalate

Concen: 2.593 ng

RT: 21.32 min Scan# 1646

Delta R.T. -0.00 min

Lab File: BE093509.D

Acq: 21 Jul 2017 14:57

Instrument :

BNA_E

ClientSampleId :


Tgt Ion:149 Resp: 330235

Ion Ratio Lower Upper

149 100

167 6.8 4.8 7.2




279 0.9 0.8 1.2



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E



#33

Indeno(1,2,3-cd)pyrene

Concen: 1.924 ng

RT: 26.55 min Scan# 2698

Delta R.T. 0.00 min

Lab File: BE093509.D

Acq: 21 Jul 2017 14:57


Tgt Ion:276 Resp: 156226

Ion Ratio Lower Upper

276 100

138 26.5 27.4 41.2#

277 25.1 20.1 30.1



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E

