

Data Path : Z:\HPCHEM1\BNA E\Data\BE072117\
 Data File : BE093511.D
 Acq On : 21 Jul 2017 16:12
 Operator : SJ/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE072117

Quant Time: Jul 21 17:20:38 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE072117.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 21 16:13:20 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
2	1,4-Dioxane	0.664	0.619	6.8	95	0.00
3	n-Nitrosodimethylamine	0.657	0.658	-0.2	109	0.00
4 S	2-Fluorophenol	1.388	1.372	1.2	101	0.00
5 S	Phenol-d6	1.945	1.849	4.9	100	0.00
6	bis(2-Chloroethyl)ether	1.548	1.555	-0.5	103	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	102	0.00
8 S	Nitrobenzene-d5	0.318	0.303	4.7	102	0.00
9	Naphthalene	4.659	4.507	3.3	102	0.00
10	Hexachlorobutadiene	0.181	0.177	2.2	102	-0.02
11 SURR	2-Methylnaphthalene-d10	0.647	0.627	3.1	101	0.00
12	2-Methylnaphthalene	0.781	0.747	4.4	101	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	103	0.00
14 S	2,4,6-Tribromophenol	0.145	0.129	11.0	88	0.00
15 S	2-Fluorobiphenyl	1.574	1.548	1.7	99	0.00
16	Acenaphthylene	2.014	1.881	6.6	100	0.00
17	Acenaphthene	1.368	1.314	3.9	99	0.00
18	Fluorene	1.926	1.841	4.4	98	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	109	0.00
20	4-Bromophenyl-phenylether	0.156	0.120	23.1	96	0.00
21	Hexachlorobenzene	0.150	0.121	19.3	97	0.00
22	Pentachlorophenol	0.088	0.061	30.7#	96	0.00
23	Phenanthrene	0.907	0.836	7.8	105	0.00
24	Anthracene	1.387	1.306	5.8	107	0.00
25 SURR	Fluoranthene-d10	6.369	5.545	12.9	98	0.00
26	Fluoranthene	1.935	1.664	14.0	98	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	99	0.00
28	Pvrene	1.296	1.284	0.9	98	0.00
29 S	Terphenyl-d14	0.728	0.726	0.3	98	0.00
30	Benzo(a)anthracene	1.286	1.243	3.3	97	0.00
31	Chrysene	1.275	1.282	-0.5	99	0.00
32	Bis(2-ethylhexyl)phthalate	2.788	2.255	19.1	87	0.00
33	Indeno(1,2,3-cd)pyrene	1.206	1.173	2.7	98	0.00
34 I	Perylene-d12	1.000	1.000	0.0	100	0.00
35	Benzo(b)fluoranthene	1.405	1.383	1.6	98	0.00
36	Benzo(k)fluoranthene	1.388	1.367	1.5	100	0.00
37 C	Benzo(a)pyrene	1.306	1.246	4.6	97	0.00
38	Dibenzo(a,h)anthracene	1.235	1.201	2.8	99	0.00
39	Benzo(a,h,i)perylene	1.240	1.205	2.8	97	0.00

Data Path : Z:\HPCHEM1\BNA E\Data\BE072117\
Data File : BE093511.D
Acq On : 21 Jul 2017 16:12
Operator : SJ/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
ICVBE072117

Quant Time: Jul 21 17:20:38 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE072117.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 21 16:13:20 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			