

Data Path : Z:\HPCHEM1\BNA E\Data\BE072117\
 Data File : BE093509.D
 Acq On : 21 Jul 2017 14:57
 Operator : SJ/JU
 Sample : SSTDICC1.6
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC1.6

Manual Integrations
 APPROVED

Sohil
 7/24/2017 5:09:32 PM

Quant Time: Jul 21 16:05:41 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE072117.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 21 15:04:06 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.931	152	2829	0.40	ng	0.00
7) Naphthalene-d8	10.707	136	13983	0.40	ng	0.00
13) Acenaphthene-d10	14.525	164	8881	0.40	ng	0.00
19) Phenanthrene-d10	17.235	188	22422	0.40	ng	0.00
27) Chrysene-d12	21.415	240	32759	0.40	ng	0.00
34) Perylene-d12	23.911	264	27006	0.40	ng	0.00

System Monitoring Compounds						
4) 2-Fluorophenol	5.513	112	15076	1.77	ng	0.00
5) Phenol-d6	7.100	99	21401	1.70	ng	0.00
8) Nitrobenzene-d5	9.061	82	17698	1.98	ng	0.00
11) 2-Methylnaphthalene-d10	12.295	152	36847	1.65	ng	0.00
14) 2,4,6-Tribromophenol	15.997	330	4507	1.61	ng	0.00
15) 2-Fluorobiphenyl	13.162	172	53738	1.56	ng	0.00
25) Fluoranthene-d10	19.276	212	549302	2.03	ng	0.00
29) Terphenyl-d14	19.860	244	95758	1.64	ng	0.00

Target Compounds						Qvalue
2) 1,4-Dioxane	3.436	88	6957	1.16	ng	98
3) n-Nitrosodimethylamine	3.738	42	7472	2.38	ng	# 88
6) bis(2-Chloroethyl)ether	7.339	93	17516	1.85	ng	# 98
9) Naphthalene	10.757	128	254415	1.74	ng	99
10) Hexachlorobutadiene	11.042	225	9912	1.78	ng	98
12) 2-Methylnaphthalene	12.367	142	44146	1.79	ng	96
16) Acenaphthylene	14.243	152	73498	1.99	ng	# 87
17) Acenaphthene	14.584	154	47994	1.77	ng	99
18) Fluorene	15.577	166	67634	1.82	ng	# 88
20) 4-Bromophenyl-phenylether	16.453	248	13562	1.10	ng	# 98
21) Hexachlorobenzene	16.583	284	12638	0.95	ng	# 100
22) Pentachlorophenol	16.909	266	8351	4.23	ng	94
23) Phenanthrene	17.268	178	77974m	0.96	ng	
24) Anthracene	17.366	178	121253m	2.43	ng	
26) Fluoranthene	19.305	202	166101	2.17	ng	98
28) Pyrene	19.660	202	170252	1.85	ng	# 98
30) Benzo(a)anthracene	21.393	228	166391	1.88	ng	94
31) Chrysene	21.449	228	166855	1.80	ng	95
32) Bis(2-ethylhexyl)phtha...	21.315	149	330235	2.59	ng	98
33) Indeno(1,2,3-cd)pyrene	26.550	276	156226	1.92	ng	# 92
35) Benzo(b)fluoranthene	23.146	252	154320	1.77	ng	# 93
36) Benzo(k)fluoranthene	23.192	252	154329	1.79	ng	# 91
37) Benzo(a)pyrene	23.802	252	142360	1.86	ng	# 91
38) Dibenzo(a,h)anthracene	26.578	278	135720	1.83	ng	# 87
39) Benzo(g,h,i)perylene	27.366	276	134970	1.80	ng	# 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA E\Data\BE072117\
Data File : BE093509.D
Acq On : 21 Jul 2017 14:57
Operator : SJ/JU
Sample : SSTDICC1.6
Misc :
ALS Vial : 6 Sample Multiplier: 1


Instrument :
BNA_E
ClientSampleId :
SSTDICC1.6

Manual Integrations
APPROVED

Sohil
7/24/2017 5:09:32 PM

Quant Time: Jul 21 16:05:41 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE072117.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 21 15:04:06 2017
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.931 min Scan# 40
 Delta R.T. 0.001 min
 Lab File: BE093509.D
 Acq: 21 Jul 2017 14:57

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC1.6

Tot Ion:152 Resp: 2829
 Ion Ratio Lower Upper
 152 100
 150 152.9 125.1 187.7
 115 67.4 55.7 83.5




Manual Integrations
 APPROVED


Sohil
 7/24/2017 5:09:32 PM



#2
 1,4-Dioxane
 Concen: 1.16 ng
 RT: 3.436 min Scan# 46
 Delta R.T. 0.001 min
 Lab File: BE093509.D
 Acq: 21 Jul 2017 14:57

Tgt Ion: 88 Resp: 6957
 Ion Ratio Lower Upper
 88 100
 58 79.8 62.2 93.4
 43 28.0 23.2 34.8









#3
n-Nitrosodimethylamine
Concen: 2.38 ng
RT: 3.738 min Scan# 70
Delta R.T. 0.001 min
Lab File: BE093509.D
Acq: 21 Jul 2017 14:57

Instrument :
BNA_E
ClientSampleId :
SSTDICC1.6

Tot Ion	Ratio	Lower	Upper
42	100		
74	137.9	99.1	148.7
44	11.4	7.4	11.2#




Manual Integrations
APPROVED




Sohil
7/24/2017 5:09:32 PM



#4
2-Fluorophenol
Concen: 1.77 ng
RT: 5.513 min Scan# 211
Delta R.T. 0.001 min
Lab File: BE093509.D
Acq: 21 Jul 2017 14:57

Tgt Ion	Ratio	Lower	Upper
112	100		
64	67.6	56.4	84.6
63	136.7	29.3	43.9#









#5
Phenol-d6
Concen: 1.70 ng
RT: 7.100 min Scan# 337
Delta R.T. 0.001 min
Lab File: BE093509.D
Acq: 21 Jul 2017 14:57

Instrument :
BNA_E
ClientSampled :
SSTDICC1.6

Tot Ion	Ratio	Resp Lower	Upper
99	100		
42	18.7	0.0	0.0#
71	33.3	0.0	0.0#


Manual Integrations
APPROVED


Sohil
7/24/2017 5:09:32 PM



#6
bis(2-Chloroethyl)ether
Concen: 1.85 ng
RT: 7.339 min Scan# 356
Delta R.T. 0.001 min
Lab File: BE093509.D
Acq: 21 Jul 2017 14:57

Tgt Ion	Ratio	Resp Lower	Upper
93	100		
63	336.6	0.0	0.0#
95	32.7	25.4	38.0





#7

Naphthalene-d8

Concen: 0.40 ng

RT: 10.707 min Scan# 57

Delta R.T. 0.000 min

Lab File: BE093509.D

Acq: 21 Jul 2017 14:57

Instrument :

BNA_E





ClientSampleId :

SSTDICC1.6

Tot Ion	136	Resp	13983
Ion Ratio	Lower	Upper	
136	100		
137	11.2	9.8	14.8
54	44.0	10.3	15.5#
68	8.8	9.0	13.4#

Manual Integrations
APPROVED

Sohil
7/24/2017 5:09:32 PM



#8

Nitrobenzene-d5

Concen: 1.98 ng




RT: 9.061 min Scan# 479


Delta R.T. 0.000 min

Lab File: BE093509.D

Acq: 21 Jul 2017 14:57

Tgt Ion	82	Resp	17698
Ion Ratio	Lower	Upper	
82	100		
128	96.2	17.0	25.4#
54	239.0	53.8	80.6#





#9

Naphthalene

Concen: 1.74 ng

RT: 10.757 min Scan# 580

Delta R.T. 0.000 min

Lab File: BE093509.D

Acq: 21 Jul 2017 14:57

Instrument :

BNA_E





ClientSampled :

SSTDICC1.6

Tot Ion:	128	Resp:	254415
Ion Ratio	Lower	Upper	
128	100		
129	2.8	2.4	3.6
127	3.3	2.4	3.6

Manual Integrations
APPROVED

Sohil
7/24/2017 5:09:32 PM



#10

Hexachlorobutadiene

Concen: 1.78 ng




RT: 11.042 min Scan# 597


Delta R.T. -0.016 min

Lab File: BE093509.D

Acq: 21 Jul 2017 14:57

Tgt Ion:	225	Resp:	9912
Ion Ratio	Lower	Upper	
225	100		
223	63.3	49.7	74.5
227	65.0	50.3	75.5









#11
2-Methylnaphthalene-d10
Concen: 1.65 ng
RT: 12.295 min Scan# 826
Delta R.T. -0.001 min
Lab File: BE093509.D
Acq: 21 Jul 2017 14:57

Instrument :
BNA_E
ClientSampleId :
SSTDICC1.6

Tot Ion	Ratio	Resp	Lower	Upper
152	100	36847		
151	19.0	15.1	22.7	




Manual Integrations
APPROVED


Sohil
7/24/2017 5:09:32 PM



#12
2-Methylnaphthalene
Concen: 1.79 ng
RT: 12.367 min Scan# 845
Delta R.T. -0.001 min
Lab File: BE093509.D
Acq: 21 Jul 2017 14:57

Tgt Ion	Ratio	Resp	Lower	Upper
142	100	44146		
141	88.6	72.6	108.8	
115	36.0	32.2	48.2	









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.525 min Scan# 1112
Delta R.T. -0.001 min
Lab File: BE093509.D
Acq: 21 Jul 2017 14:57

Instrument :
BNA_E
ClientSampleId :
SSTDICC1.6

Tot Ion	Ratio	Lower	Upper
164	100		
162	105.1	84.6	127.0
160	48.2	41.8	62.6




Manual Integrations
APPROVED




Sohil
7/24/2017 5:09:32 PM



#14
2,4,6-Tribromophenol
Concen: 1.61 ng
RT: 15.997 min Scan# 1205
Delta R.T. -0.001 min
Lab File: BE093509.D
Acq: 21 Jul 2017 14:57

Tgt Ion	Ratio	Lower	Upper
330	100		
332	90.1	77.7	116.5
141	49.8	56.2	84.4#









#15
2-Fluorobiphenyl
Concen: 1.56 ng
RT: 13.162 min Scan# 1015
Delta R.T. -0.001 min
Lab File: BE093509.D
Acq: 21 Jul 2017 14:57

Instrument :
BNA_E
ClientSampleId :
SSTDICC1.6

Tot Ion	Ratio	Lower	Upper
172	100		
171	33.3	29.8	44.6
170	21.1	20.7	31.1


Manual Integrations
APPROVED


Sohil
7/24/2017 5:09:32 PM



#16
Acenaphthylene
Concen: 1.99 ng
RT: 14.243 min Scan# 1093
Delta R.T. -0.001 min
Lab File: BE093509.D
Acq: 21 Jul 2017 14:57

Tgt Ion	Ratio	Lower	Upper
152	100		
151	19.6	4.0	6.0#
153	13.1	10.3	15.5









#17
Acenaphthene
Concen: 1.77 ng
RT: 14.584 min Scan# 1116
Delta R.T. -0.001 min
Lab File: BE093509.D
Acq: 21 Jul 2017 14:57

Instrument :
BNA_E
ClientSampleId :
SSTDICC1.6

Tot Ion	Ratio	Lower	Upper
154	100		
153	114.9	91.2	136.8
152	54.9	44.8	67.2




Manual Integrations
APPROVED




Sohil
7/24/2017 5:09:32 PM



#18
Fluorene
Concen: 1.82 ng
RT: 15.577 min Scan# 1183
Delta R.T. 0.014 min
Lab File: BE093509.D
Acq: 21 Jul 2017 14:57

Tgt Ion	Ratio	Lower	Upper
166	100		
165	98.3	78.6	117.8
167	53.9	11.1	16.7#





#19

Phenanthrene-d10

Concen: 0.40 ng

RT: 17.235 min Scan# 12

Delta R.T. -0.001 min

Lab File: BE093509.D

Acq: 21 Jul 2017 14:57

Instrument :

BNA_E





ClientSampleId :

SSTDICC1.6

Tot Ion:	188	Resp:	22422
Ion Ratio	Lower	Upper	
188	100		
94	17.8	11.3	16.9#
80	18.6	11.7	17.5#

Manual Integrations
APPROVED

Sohil
7/24/2017 5:09:32 PM



#20

4-Bromophenyl-phenylether

Concen: 1.10 ng


RT: 16.453 min Scan# 1219


Delta R.T. 0.000 min

Lab File: BE093509.D

Acq: 21 Jul 2017 14:57

Tgt Ion:	248	Resp:	13562
Ion Ratio	Lower	Upper	
248	100		
250	158.9	128.0	192.0
141	8.2	12.0	18.0#









#21
Hexachlorobenzene
Concen: 0.95 ng
RT: 16.583 min Scan# 1223
Delta R.T. -0.001 min
Lab File: BE093509.D
Acq: 21 Jul 2017 14:57

Instrument :
BNA_E
ClientSampleId :
SSTDICC1.6

Tot Ion:	284	Resp:	12638
Ion Ratio	Lower	Upper	
284	100		
142	0.0	0.0	0.0
249	0.0	0.0	0.0




Manual Integrations
APPROVED




Sohil
7/24/2017 5:09:32 PM



#22
Pentachlorophenol
Concen: 4.23 ng
RT: 16.909 min Scan# 1233
Delta R.T. -0.001 min
Lab File: BE093509.D
Acq: 21 Jul 2017 14:57

Tgt Ion:	266	Resp:	8351
Ion Ratio	Lower	Upper	
266	100		
264	61.6	56.0	84.0
268	63.4	52.0	78.0





#23

Phenanthrene

Concen: 0.96 ng m

RT: 17.268 min Scan# 1244

Delta R.T. -0.001 min

Lab File: BE093509.D

Acq: 21 Jul 2017 14:57

Instrument :

BNA_E





ClientSampled :

SSTDICC1.6

Tot Ion	Ratio	Lower	Upper
178	100		
176	0.0	15.0	22.4#
179	0.0	12.7	19.1#

Manual Integrations
APPROVED

Sohil
7/24/2017 5:09:32 PM



#24

Anthracene

Concen: 2.43 ng m


RT: 17.366 min Scan# 1247


Delta R.T. -0.001 min

Lab File: BE093509.D

Acq: 21 Jul 2017 14:57

Tgt Ion	Ratio	Lower	Upper
178	100		
176	34.0	14.6	22.0#
179	22.7	12.0	18.0#









#25
Fluoranthene-d10
Concen: 2.03 ng
RT: 19.276 min Scan# 1432
Delta R.T. 0.001 min
Lab File: BE093509.D
Acq: 21 Jul 2017 14:57

Instrument :
BNA_E
ClientSampleId :
SSTDICC1.6

Tot Ion:	212	Resp:	549302
Ion Ratio		Lower	Upper
212	100		
106	3.3	2.4	3.6
104	1.8	1.6	2.4




Manual Integrations
APPROVED


Sohil
7/24/2017 5:09:32 PM



#26
Fluoranthene
Concen: 2.17 ng
RT: 19.305 min Scan# 1436
Delta R.T. 0.001 min
Lab File: BE093509.D
Acq: 21 Jul 2017 14:57

Tgt Ion:	202	Resp:	166101
Ion Ratio		Lower	Upper
202	100		
101	11.3	9.6	14.4
203	17.3	14.4	21.6









#27
Chrysene-d12
Concen: 0.40 ng
RT: 21.415 min Scan# 1655
Delta R.T. -0.000 min
Lab File: BE093509.D
Acq: 21 Jul 2017 14:57

Instrument :
BNA_E
ClientSampleId :
SSTDICC1.6

Tot Ion	Ratio	Lower	Upper
240	100		
120	5.9	4.6	7.0
236	24.6	20.8	31.2




Manual Integrations
APPROVED


Sohil
7/24/2017 5:09:32 PM



#28
Pyrene
Concen: 1.85 ng
RT: 19.660 min Scan# 1484
Delta R.T. -0.006 min
Lab File: BE093509.D
Acq: 21 Jul 2017 14:57

Tgt Ion	Ratio	Lower	Upper
202	100		
200	20.9	0.0	0.0#
203	17.8	13.7	20.5





#29

Terphenyl-d14

Concen: 1.64 ng

RT: 19.860 min Scan# 1512

Delta R.T. -0.006 min

Lab File: BE093509.D


Acq: 21 Jul 2017 14:57

Instrument :

BNA_E

ClientSampleId :

SSTDICC1.6



Tot Ion:244 Resp: 95758

Ion Ratio Lower Upper




244 100

212 32.8 10.6 16.0#

122 12.9 15.4 23.0#

Manual Integrations
APPROVED

Sohil
7/24/2017 5:09:32 PM



#30

Benzo(a)anthracene

Concen: 1.88 ng

RT: 21.393 min Scan# 1653

Delta R.T. -0.000 min

Lab File: BE093509.D

Acq: 21 Jul 2017 14:57




Tgt Ion:228 Resp: 166391


Ion Ratio Lower Upper

228 100

226 26.8 19.7 29.5

229 20.0 19.2 28.8









#31
Chrysene
Concen: 1.80 ng
RT: 21.449 min Scan# 1658
Delta R.T. -0.000 min
Lab File: BE093509.D
Acq: 21 Jul 2017 14:57

Instrument :
BNA_E
ClientSampleId :
SSTDICC1.6

Tot Ion	Ratio	Resp	Lower	Upper
228	100	166855		
226	28.9		21.5	32.3
229	20.2		18.6	28.0




Manual Integrations
APPROVED


Sohil
7/24/2017 5:09:32 PM



#32
Bis(2-ethylhexyl)phthalate
Concen: 2.59 ng
RT: 21.315 min Scan# 1646
Delta R.T. -0.000 min
Lab File: BE093509.D
Acq: 21 Jul 2017 14:57

Tgt Ion	Ratio	Resp	Lower	Upper
149	100	330235		
167	6.8		4.8	7.2
279	0.9		0.8	1.2









#33
Indeno(1,2,3-cd)pyrene
Concen: 1.92 ng
RT: 26.550 min Scan# 2698
Delta R.T. 0.000 min
Lab File: BE093509.D
Acq: 21 Jul 2017 14:57

Instrument :
BNA_E
ClientSampleId :
SSTDICC1.6

Tot Ion	Ratio	Lower	Upper
276	100		
138	26.5	27.4	41.2#
277	25.1	20.1	30.1




Manual Integrations
APPROVED


Sohil
7/24/2017 5:09:32 PM



#34
Perylene-d12
Concen: 0.40 ng
RT: 23.911 min Scan# 2122
Delta R.T. -0.006 min
Lab File: BE093509.D
Acq: 21 Jul 2017 14:57

Tgt Ion	Ratio	Lower	Upper
264	100		
260	24.8	21.0	31.4
265	32.3	24.6	36.8





#35

Benzo(b)fluoranthene

Concen: 1.77 ng

RT: 23.146 min Scan# 1953

Delta R.T. 0.004 min

Lab File: BE093509.D

Acq: 21 Jul 2017 14:57

Instrument :

BNA_E





ClientSampleId :

SSTDICC1.6

Tot Ion:	252	Resp:	154320
Ion Ratio	Lower	Upper	
252	100		
253	22.3	18.5	27.7
125	10.4	13.4	20.0#

Manual Integrations
APPROVED

Sohil
7/24/2017 5:09:32 PM



#36

Benzo(k)fluoranthene

Concen: 1.79 ng




RT: 23.192 min Scan# 1964


Delta R.T. -0.001 min

Lab File: BE093509.D

Acq: 21 Jul 2017 14:57

Tgt Ion:	252	Resp:	154329
Ion Ratio	Lower	Upper	
252	100		
253	21.7	20.3	30.5
125	10.6	12.2	18.4#









#37
Benzo(a)pyrene
Concen: 1.86 ng
RT: 23.802 min Scan# 2097
Delta R.T. 0.004 min
Lab File: BE093509.D
Acq: 21 Jul 2017 14:57

Instrument :
BNA_E
ClientSampleId :
SSTDICC1.6

Tot Ion	Ratio	Resp Lower	Upper
252	100		
253	22.1	19.9	29.9
125	11.7	14.6	21.8




Manual Integrations
APPROVED


Sohil
7/24/2017 5:09:32 PM



#38
Dibenzo(a,h)anthracene
Concen: 1.83 ng
RT: 26.578 min Scan# 2704
Delta R.T. 0.005 min
Lab File: BE093509.D
Acq: 21 Jul 2017 14:57

Tgt Ion	Ratio	Resp Lower	Upper
278	100		
139	17.0	21.4	32.0
279	24.0	22.2	33.2





#39

Benzo(a,h,i)perylene

Concen: 1.80 ng

RT: 27.366 min Scan# 28

Delta R.T. 0.000 min

Lab File: BE093509.D

Acq: 21 Jul 2017 14:57

Instrument :

BNA_E

ClientSampleId :

SSTDICC1.6

Tot Ion:	276	Resp:	134970
Ion Ratio	Lower	Upper	
276	100		
277	24.0	21.2	31.8
138	22.8	24.5	36.7#

Manual Integrations
APPROVED

Sohil
7/24/2017 5:09:32 PM

