




Data Path : Z:\HPCHEM1\BNA E\DATA\BE072215\
Data File : BE090211.D
Acq On : 21 Jul 2015 22:15
Operator : TP/IZ
Sample : MDL-02-S
Misc : MDL-SIM-SOIL-0.07PPM
ALS Vial : 10 Sample Multiplier: 1

Quant Time: Jul 22 04:57:48 2015
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM02.2-EPA-SIM-BE072215.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 22 04:13:39 2015
Response via : Initial Calibration



TIC: BE090211.D

(22) Benzo(k)fluoranthene

22.767min (+0.004) 0.04ng/ul m

response 255

Ion	Exp%	Act%
252.00	100	100
253.00	24.40	68.00#
125.00	14.90	66.40#
0.00	0.00	0.00