

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE072215\
 Data File : BE090208.D
 Acq On : 21 Jul 2015 20:24
 Operator : TP/IZ
 Sample : SSTD3.223
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTD3.223

Quant Time: Jul 22 03:48:56 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM02.2-EPA-SIM-BE072215.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 22 03:03:48 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.60	152	1009	0.40	ng/ul	0.00
2) Naphthalene-d8	10.36	136	5006	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.21	164	2908	0.40	ng/ul	0.00
10) Phenanthrene-d10	16.94	188	7457	0.40	ng/ul	0.00
16) Chrysene-d12	21.11	240	8959	0.40	ng/ul	0.00
20) Perylene-d12	23.43	264	6286	0.40	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units
4) 2-Methylnaphthalene-d10	0.00	152	0d	0.00	ng/ul
14) Fluoranthene-d10	0.00	212	0d	0.00	ng/ul


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
11) Pentachlorophenol	16.60	266	1952	2.75	ng/ul#	90




(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE072215\
Data File : BE090208.D
Acq On : 21 Jul 2015 20:24
Operator : TP/IZ
Sample : SSTD3.223
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTD3.223

Quant Time: Jul 22 03:48:56 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM02.2-EPA-SIM-BE072215.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 22 03:03:48 2015
Response via : Initial Calibration





#11

Pentachlorophenol

Concen: 2.75 ng/ul

RT: 16.60 min Scan# 2186

Delta R.T. -0.01 min

Lab File: BE090208.D

Acq: 21 Jul 2015 20:24

Instrument :

BNA_E

ClientSampleId :

SSTD3.223

Tgt Ion: 266 Resp: 1952

Ion Ratio Lower Upper

266 100

264 63.3 51.6 77.4

268 65.4 41.3 61.9#

Abundance Ion 266.00 (265.70 to 266.70): E

Ion 264.00 (263.70 to 264.70): E

Ion 268.00 (267.70 to 268.70): E

