

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE072216\
 Data File : BE092015.D
 Acq On : 22 Jul 2016 20:46
 Operator : UM/SJ
 Sample : H4120-06
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 DUP-01

Manual Integrations

sohil
 7/25/2016 6:46:13 PM

Quant Time: Jul 23 02:00:32 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE072216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 22 15:39:27 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.20	152	25011	5.00	ng	0.00
8) Naphthalene-d8	10.99	136	124318	5.00	ng	0.00
15) Acenaphthene-d10	14.86	164	82624	5.00	ng	0.00
24) Phenanthrene-d10	17.59	188	191344	5.00	ng	0.00
31) Chrysene-d12	21.76	240	251663	5.00	ng	0.00
40) Perylene-d12	24.16	264	202105	5.00	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.72	112	15970	3.29	ng	0.00
5) Phenol-d6	7.35	99	15953	1.94	ng	-0.02
9) Nitrobenzene-d5	9.37	82	27059	3.62	ng	-0.02
16) 2,4,6-Tribromophenol	16.35	330	18545	6.50	ng	0.00
17) 2-Fluorobiphenyl	13.47	172	77211	3.55	ng	-0.01
34) Terphenyl-d14	20.20	244	133990	5.30	ng	-0.02
Target Compounds						
22) Fluorene	15.90	166	4517	0.18	ng	# 60
28) Phenanthrene	17.63	178	1759	0.04	ng	# 79
29) Anthracene	17.72	178	1786	0.04	ng	# 84
30) Fluoranthene	19.63	202	15967	0.08	ng	# 67
33) Pyrene	19.99	202	20017	0.11	ng	# 70
35) Benzo(a)anthracene	21.73	228	2319m	0.04	ng	
38) Bis(2-ethylhexyl)phthalate	21.64	149	9826	0.27	ng	# 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE072216\
Data File : BE092015.D
Acq On : 22 Jul 2016 20:46
Operator : UM/SJ
Sample : H4120-06
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
DUP-01

Manual Integrations

sohil
7/25/2016 6:46:13 PM

Quant Time: Jul 23 02:00:32 2016
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE072216.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 22 15:39:27 2016
Response via : Initial Calibration

