

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE072417\
 Data File : BE093534.D
 Acq On : 24 Jul 2017 12:32
 Operator : SJ/JU
 Sample : SSTD3.234
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :

Quant Time: Jul 24 13:09:31 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE072417.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 24 11:37:06 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.93	152	2926	0.40	ng/ul	0.00
2) Naphthalene-d8	10.71	136	15065	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.54	164	9820	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.26	188	25381	0.40	ng/ul	0.00
16) Chrysene-d12	21.42	240	30857	0.40	ng/ul	0.00
20) Perylene-d12	23.92	264	26336	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	0.00	152	0d	0.00	ng/ul	
14) Fluoranthene-d10	0.00	212	0d	0.00	ng/ul	
Target Compounds						
11) Pentachlorophenol	16.93	266	15085	4.006	ng/ul	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE072417\
Data File : BE093534.D
Acq On : 24 Jul 2017 12:32
Operator : SJ/JU
Sample : SST3.234
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :

Quant Time: Jul 24 13:09:31 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE072417.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 24 11:37:06 2017
Response via : Initial Calibration

