

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE072417\
 Data File : BE093538.D
 Acq On : 24 Jul 2017 15:01
 Operator : SJ/JU
 Sample : PB100726BL
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SBLK26

Manual Integrations
 APPROVED

Sohil
 7/25/2017 5:32:32 PM

Quant Time: Jul 24 17:29:09 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE072417.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 24 13:14:05 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.92	152	3131	0.40	ng/ul	0.00
2) Naphthalene-d8	10.71	136	16835	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.53	164	11599	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.26	188	28781m	0.40	ng/ul	0.00
16) Chrysene-d12	21.42	240	30770	0.40	ng/ul	0.00
20) Perylene-d12	23.92	264	24635m	0.40	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Methylnaphthalene-d10	12.29	152	8690	0.32	ng/ul	0.00
14) Fluoranthene-d10	19.28	212	28493	0.31	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE072417\
Data File : BE093538.D
Acq On : 24 Jul 2017 15:01
Operator : SJ/JU
Sample : PB100726BL
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SBLK26

Manual Integrations
APPROVED

Sohil
7/25/2017 5:32:32 PM

Quant Time: Jul 24 17:29:09 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE072417.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 24 13:14:05 2017
Response via : Initial Calibration

