

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE072417\
 Data File : BE093571.D
 Acq On : 25 Jul 2017 11:15
 Operator : SJ/JU
 Sample : I4209-11
 Misc :
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 C0B14

Manual Integrations
 APPROVED

Sohil
 7/25/2017 5:34:19 PM

Quant Time: Jul 25 13:06:49 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE072417.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 25 06:10:24 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.92	152	2302	0.40	ng/ul	0.00
2) Naphthalene-d8	10.71	136	12290	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.53	164	8416	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.26	188	21487m	0.40	ng/ul	0.00
16) Chrysene-d12	21.41	240	23610	0.40	ng/ul	0.00
20) Perylene-d12	23.91	264	19295m	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.29	152	6201	0.31	ng/ul	0.00
14) Fluoranthene-d10	19.28	212	21016	0.31	ng/ul	0.00
Target Compounds					Qvalue	
3) Naphthalene	10.76	128	2299	0.077	ng/ul#	94
12) Phenanthrene	17.29	178	1903	0.033	ng/ul#	91
15) Fluoranthene	19.31	202	4651	0.064	ng/ul	85
17) Pyrene	19.67	202	5258	0.076	ng/ul#	86
18) Benzo(a)anthracene	21.40	228	4300	0.063	ng/ul#	88
19) Chrysene	21.45	228	5279	0.080	ng/ul	95
21) Benzo(b)fluoranthene	23.15	252	7463	0.125	ng/ul	91
22) Benzo(k)fluoranthene	23.19	252	2987m	0.051	ng/ul	
23) Benzo(a)pyrene	23.81	252	6125m	0.108	ng/ul	
24) Indeno(1,2,3-cd)pyrene	26.55	276	3820m	0.061	ng/ul	
26) Benzo(g,h,i)perylene	27.37	276	3858m	0.073	ng/ul	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE072417\
Data File : BE093571.D
Acq On : 25 Jul 2017 11:15
Operator : SJ/JU
Sample : I4209-11
Misc :
ALS Vial : 42 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
C0B14

Manual Integrations
APPROVED

Sohil
7/25/2017 5:34:19 PM

Quant Time: Jul 25 13:06:49 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE072417.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 25 06:10:24 2017
Response via : Initial Calibration

