

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE072417\
 Data File : BE093540.D
 Acq On : 24 Jul 2017 16:13
 Operator : SJ/JU
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.436

Quant Time: Jul 24 17:38:07 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE072417.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 24 13:14:05 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	93	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	91	0.00
3	Naphthalene	0.973	0.981	-0.8	93	0.00
4 SURR	2-Methylnaphthalene-d10	0.648	0.660	-1.9	94	0.00
5	2-Methylnaphthalene	0.710	0.711	-0.1	93	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	92	0.00
7	Acenaphthylene	1.673	1.666	0.4	95	0.00
8 C	Acenaphthene	1.316	1.314	0.2	94	0.00
9	Fluorene	1.664	1.644	1.2	93	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	93	0.00
11	Pentachlorophenol	0.066	0.051	22.7#	80	0.00
12	Phenanthrene	1.089	1.081	0.7	96	0.00
13	Anthracene	1.061	1.016	4.2	93	0.00
14 SURR	Fluoranthene-d10	1.263	1.195	5.4	91	0.00
15 C	Fluoranthene	1.362	1.275	6.4	91	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	91	0.00
17	Pyrene	1.168	1.140	2.4	91	0.00
18	Benzo(a)anthracene	1.160	1.109	4.4	90	0.00
19	Chrysene	1.112	1.095	1.5	92	0.00
20 I	Perylene-d12	1.000	1.000	0.0	90	0.00
21	Benzo(b)fluoranthene	1.237	1.211	2.1	92	0.00
22	Benzo(k)fluoranthene	1.220	1.176	3.6	90	0.00
23 C	Benzo(a)pyrene	1.177	1.136	3.5	91	0.00
24	Indeno(1,2,3-cd)pyrene	1.299	1.260	3.0	92	0.00
25	Dibenzo(a,h)anthracene	1.086	1.064	2.0	93	0.00
26	Benzo(g,h,i)perylene	1.095	1.056	3.6	91	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0