

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE072418\
 Data File : BE097089.D
 Acq On : 25 Jul 2018 1:58
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 100 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :

Quant Time: Jul 25 07:38:56 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE072418.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 24 13:24:10 2018
 Response via : Initial Calibration

Internal Standards	R.T.	OIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.39	152	1323	0.40	ng	0.00
7) Naphthalene-d8	10.15	136	6727	0.40	ng	0.00
13) Acenaphthene-d10	14.01	164	4297	0.40	ng	0.00
19) Phenanthrene-d10	16.77	188	11502	0.40	ng	0.00
27) Chrysene-d12	20.98	240	10878	0.40	ng	0.00
34) Perylene-d12	23.21	264	9716	0.40	ng	0.00

System Monitoring Compounds

4) 2-Fluorophenol	5.05	112	1527	0.41	ng	0.00
5) Phenol-d6	6.60	99	2356	0.43	ng	0.00
8) Nitrobenzene-d5	8.52	82	2393	0.41	ng	0.00
11) 2-Methylnaphthalene-d10	11.73	152	4024	0.40	ng	0.00
14) 2,4,6-Tribromophenol	15.52	330	580	0.41	ng	0.00
15) 2-Fluorobiphenyl	12.63	172	5989	0.40	ng	0.00
25) Fluoranthene-d10	18.81	212	46129	0.40	ng	0.00
29) Terphenyl-d14	19.43	244	8459	0.40	ng	0.00

Target Compounds

						Ovalue
2) 1,4-Dioxane	3.10	88	844	0.401	ng	# 90
3) n-Nitrosodimethylamine	3.38	42	943	0.401	ng	# 93
6) bis(2-Chloroethyl)ether	6.84	93	2097	0.421	ng	# 87
9) Naphthalene	10.20	128	24232	0.401	ng	# 100
10) Hexachlorobutadiene	10.48	225	1077	0.398	ng	# 99
12) 2-Methylnaphthalene	11.80	142	4163	0.402	ng	# 99
16) Acenaphthylene	13.73	152	7862	0.408	ng	# 100
17) Acenaphthene	14.08	154	4618	0.412	ng	# 94
18) Fluorene	15.08	166	6136	0.414	ng	# 80
20) 4-Bromophenyl-phenylether	15.97	248	2232	0.399	ng	# 84
21) Hexachlorobenzene	16.08	284	2458	0.410	ng	# 84
22) Pentachlorophenol	16.43	266	584	0.416	ng	# 79
23) Phenanthrene	16.81	178	11136	0.406	ng	# 99
24) Anthracene	16.89	178	9875	0.403	ng	# 95
26) Fluoranthene	18.84	202	12372	0.402	ng	# 98
28) Pyrene	19.20	202	12316	0.413	ng	# 99
30) Benzo(a)anthracene	20.95	228	10752	0.397	ng	# 98
31) Chrysene	21.00	228	12216	0.421	ng	# 98
32) Bis(2-ethylhexyl)phthalate	20.92	149	13985	0.485	ng	# 99
33) Indeno(1,2,3-cd)pyrene	25.51	276	10894	0.415	ng	# 92
35) Benzo(b)fluoranthene	22.54	252	9601	0.389	ng	# 87
36) Benzo(k)fluoranthene	22.58	252	11099	0.386	ng	# 96
37) Benzo(a)pyrene	23.11	252	9271	0.400	ng	# 92
38) Dibenzo(a,h)anthracene	25.54	278	9001	0.389	ng	# 96
39) Benzo(a,h,i)perylene	26.20	276	9464	0.385	ng	# 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE072418\
Data File : BE097089.D
Acq On : 25 Jul 2018 1:58
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 100 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :

Quant Time: Jul 25 02:36:54 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE072418.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
OLast Update : Tue Jul 24 13:24:10 2018
Response via : Initial Calibration

