

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE072418\
 Data File : BE097096.D
 Acq On : 25 Jul 2018 6:59
 Operator : SJ/JU
 Sample : J4073-07DL 2X
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 GW04-071718DL

Quant Time: Jul 25 08:57:49 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE072418.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 24 13:24:10 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.39	152	1376	0.40	ng	0.00
7) Naphthalene-d8	10.15	136	6486	0.40	ng	0.00
13) Acenaphthene-d10	14.01	164	4003	0.40	ng	0.00
19) Phenanthrene-d10	16.77	188	11853	0.40	ng	0.00
27) Chrysene-d12	20.98	240	10778	0.40	ng	0.00
34) Perylene-d12	23.21	264	10070	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.06	112	363	0.09	ng	0.00
5) Phenol-d6	6.60	99	296	0.05	ng	0.00
8) Nitrobenzene-d5	8.53	82	1309	0.23	ng	0.01
11) 2-Methylnaphthalene-d10	11.73	152	2051	0.21	ng	0.00
14) 2,4,6-Tribromophenol	15.52	330	336	0.26	ng	0.00
15) 2-Fluorobiphenyl	12.63	172	4960	0.36	ng	0.00
25) Fluoranthene-d10	18.81	212	27391	0.23	ng	0.00
29) Terphenyl-d14	19.42	244	7985	0.39	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.10	88	3646	1.667	ng	Qvalue # 86
6) bis(2-Chloroethyl)ether	6.84	93	122	0.024	ng	84
32) Bis(2-ethylhexyl)phthalate	20.92	149	1231	0.033	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE072418\
Data File : BE097096.D
Acq On : 25 Jul 2018 6:59
Operator : SJ/JU
Sample : J4073-07DL 2X
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
GW04-071718DL

Quant Time: Jul 25 08:57:49 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE072418.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 24 13:24:10 2018
Response via : Initial Calibration

