

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE072418\
 Data File : BE097089.D
 Acq On : 25 Jul 2018 1:58
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 100 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 25 02:36:54 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE072418.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 24 13:24:10 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
2	1,4-Dioxane	0.636	0.638	-0.3	99	0.00
3	n-Nitrosodimethylamine	0.711	0.713	-0.3	103	0.00
4 S	2-Fluorophenol	1.124	1.154	-2.7	104	0.00
5 S	Phenol-d6	1.649	1.781	-8.0	111	0.00
6	bis(2-Chloroethyl)ether	1.506	1.585	-5.2	109	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	98	0.00
8 S	Nitrobenzene-d5	0.347	0.356	-2.6	106	0.00
9	Naphthalene	3.590	3.602	-0.3	103	0.00
10	Hexachlorobutadiene	0.161	0.160	0.6	101	0.00
11 SURR	2-Methylnaphthalene-d10	0.599	0.598	0.2	102	0.00
12	2-Methylnaphthalene	0.616	0.619	-0.5	103	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	99	0.00
14 S	2,4,6-Tribromophenol	0.132	0.135	-2.3	113	0.00
15 S	2-Fluorobiphenyl	1.381	1.394	-0.9	103	0.00
16	Acenaphthylene	1.795	1.830	-1.9	104	0.00
17	Acenaphthene	1.044	1.075	-3.0	105	0.00
18	Fluorene	1.379	1.428	-3.6	107	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	99	0.00
20	4-Bromophenyl-phenylether	0.194	0.194	0.0	103	0.00
21	Hexachlorobenzene	0.209	0.214	-2.4	103	0.00
22	Pentachlorophenol	0.057	0.051	10.5	110	-0.01
23	Phenanthrene	0.953	0.968	-1.6	105	0.00
24	Anthracene	0.853	0.859	-0.7	107	-0.01
25 SURR	Fluoranthene-d10	4.003	4.011	-0.2	108	0.00
26	Fluoranthene	1.070	1.076	-0.6	106	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	101	0.00
28	Pvrene	1.097	1.132	-3.2	110	0.00
29 S	Terphenyl-d14	0.768	0.778	-1.3	109	0.00
30	Benzo(a)anthracene	0.995	0.988	0.7	112	0.00
31	Chrysene	1.067	1.123	-5.2	110	-0.01
32	Bis(2-ethylhexyl)phthalate	1.243	1.286	-3.5	128	0.00
33	Indeno(1,2,3-cd)pyrene	0.966	1.001	-3.6	112	0.00
34 I	Perylene-d12	1.000	1.000	0.0	109	0.00
35	Benzo(b)fluoranthene	1.015	0.988	2.7	112	0.00
36	Benzo(k)fluoranthene	1.185	1.142	3.6	113	0.00
37 C	Benzo(a)pyrene	0.955	0.954	0.1	115	0.00
38	Dibenzo(a,h)anthracene	0.953	0.926	2.8	112	0.00
39	Benzo(a,h,i)perylene	1.011	0.974	3.7	111	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE072418\
Data File : BE097089.D
Acq On : 25 Jul 2018 1:58
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 100 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 25 02:36:54 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE072418.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 24 13:24:10 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			