

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE072418\
 Data File : BE097099.D
 Acq On : 25 Jul 2018 8:50
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 100 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 25 09:28:51 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE072418.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 24 13:24:10 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	92	0.00
2	1,4-Dioxane	0.400	0.404	-1.0	96	0.00
3	n-Nitrosodimethylamine	0.400	0.399	0.3	99	0.00
4 S	2-Fluorophenol	0.400	0.389	2.8	95	0.00
5 S	Phenol-d6	0.400	0.381	4.8	94	0.00
6	bis(2-Chloroethyl)ether	0.400	0.414	-3.5	103	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	89	0.00
8 S	Nitrobenzene-d5	0.400	0.401	-0.3	94	0.00
9	Naphthalene	0.400	0.405	-1.3	95	0.00
10	Hexachlorobutadiene	0.400	0.427	-6.7	99	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.392	2.0	91	0.00
12	2-Methylnaphthalene	0.400	0.395	1.3	92	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	87	0.00
14 S	2,4,6-Tribromophenol	0.400	0.367	8.3	88	0.00
15 S	2-Fluorobiphenyl	0.400	0.408	-2.0	91	0.00
16	Acenaphthylene	0.400	0.409	-2.2	91	0.00
17	Acenaphthene	0.400	0.408	-2.0	91	0.00
18	Fluorene	0.400	0.411	-2.7	93	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	89	0.00
20	4-Bromophenyl-phenylether	0.400	0.393	1.8	91	0.00
21	Hexachlorobenzene	0.400	0.414	-3.5	94	0.00
22	Pentachlorophenol	0.400	0.386	3.5	90	0.00
23	Phenanthrene	0.400	0.402	-0.5	93	0.00
24	Anthracene	0.400	0.396	1.0	94	0.00
25 SURR	Fluoranthene-d10	0.400	0.404	-1.0	98	0.00
26	Fluoranthene	0.400	0.409	-2.2	97	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	90	0.00
28	Pvrene	0.400	0.419	-4.7	99	0.00
29 S	Terphenyl-d14	0.400	0.410	-2.5	98	0.00
30	Benzo(a)anthracene	0.400	0.391	2.3	98	0.00
31	Chrysene	0.400	0.426	-6.5	99	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.398	0.5	92	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.394	1.5	95	0.00
34 I	Perylene-d12	0.400	0.400	0.0	96	0.00
35	Benzo(b)fluoranthene	0.400	0.377	5.8	96	0.00
36	Benzo(k)fluoranthene	0.400	0.398	0.5	103	0.00
37 C	Benzo(a)pyrene	0.400	0.379	5.3	96	0.00
38	Dibenzo(a,h)anthracene	0.400	0.373	6.8	95	0.00
39	Benzo(a,h,i)perylene	0.400	0.371	7.3	94	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE072418\
Data File : BE097099.D
Acq On : 25 Jul 2018 8:50
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 100 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 25 09:28:51 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE072418.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 24 13:24:10 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			