

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE072517\
 Data File : BE093588.D
 Acq On : 25 Jul 2017 21:45
 Operator : SJ/JU
 Sample : I4209-17
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 C0B20

Manual Integrations
 APPROVED

Sohil
 7/26/2017 5:56:05 PM

Quant Time: Jul 26 04:46:54 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE072417.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 25 06:10:24 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.92	152	2462	0.40	ng/ul	0.00
2) Naphthalene-d8	10.71	136	13277	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.53	164	8777	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.26	188	21785	0.40	ng/ul	0.00
16) Chrysene-d12	21.41	240	23587	0.40	ng/ul	0.00
20) Perylene-d12	23.91	264	20912m	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.29	152	3700	0.17	ng/ul	0.00
14) Fluoranthene-d10	19.28	212	12014	0.17	ng/ul	0.00
Target Compounds					Qvalue	
3) Naphthalene	10.76	128	1101	0.034	ng/ul#	89
12) Phenanthrene	17.29	178	3656	0.062	ng/ul#	89
15) Fluoranthene	19.30	202	9177	0.124	ng/ul	85
17) Pyrene	19.66	202	9994	0.145	ng/ul#	87
18) Benzo(a)anthracene	21.39	228	6482	0.095	ng/ul#	86
19) Chrysene	21.45	228	6293	0.096	ng/ul	96
21) Benzo(b)fluoranthene	23.14	252	8148	0.126	ng/ul	97
22) Benzo(k)fluoranthene	23.19	252	3801m	0.060	ng/ul	
23) Benzo(a)pyrene	23.80	252	6293m	0.102	ng/ul	
24) Indeno(1,2,3-cd)pyrene	26.55	276	3752m	0.055	ng/ul	
26) Benzo(g,h,i)perylene	27.36	276	2952m	0.052	ng/ul	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE072517\
Data File : BE093588.D
Acq On : 25 Jul 2017 21:45
Operator : SJ/JU
Sample : I4209-17
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
C0B20

Manual Integrations
APPROVED

Sohil
7/26/2017 5:56:05 PM

Quant Time: Jul 26 04:46:54 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE072417.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 25 06:10:24 2017
Response via : Initial Calibration

