

Data Path : Z:\HPCHEM1\BNA\_E\DATA\BE072517\  
 Data File : BE093589.D  
 Acq On : 25 Jul 2017 22:21  
 Operator : SJ/JU  
 Sample : I4209-18  
 Misc :  
 ALS Vial : 15 Sample Multiplier: 1

Instrument :  
 BNA\_E  
 ClientSampleId :  
 C0B21

Manual Integrations  
 APPROVED

Sohil  
 7/26/2017 5:56:07 PM

Quant Time: Jul 26 10:37:24 2017  
 Quant Method : Z:\HPCHEM1\BNA\_E\METHODS\SOM-EPA-SIM-BE072417.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Tue Jul 25 06:10:24 2017  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc  | Units  | Dev(Min) |
|-----------------------------|-------|------|----------|-------|--------|----------|
| 1) 1,4-Dichlorobenzene-d4   | 7.93  | 152  | 2418     | 0.40  | ng/ul  | 0.00     |
| 2) Naphthalene-d8           | 10.71 | 136  | 12833    | 0.40  | ng/ul  | 0.00     |
| 6) Acenaphthene-d10         | 14.53 | 164  | 8425     | 0.40  | ng/ul  | 0.00     |
| 10) Phenanthrene-d10        | 17.26 | 188  | 19809m   | 0.40  | ng/ul  | 0.00     |
| 16) Chrysene-d12            | 21.41 | 240  | 21081    | 0.40  | ng/ul  | 0.00     |
| 20) Perylene-d12            | 23.91 | 264  | 18795m   | 0.40  | ng/ul  | 0.00     |
| System Monitoring Compounds |       |      |          |       |        |          |
| 4) 2-Methylnaphthalene-d10  | 12.29 | 152  | 6821     | 0.33  | ng/ul  | 0.00     |
| 14) Fluoranthene-d10        | 19.28 | 212  | 20775    | 0.33  | ng/ul  | 0.00     |
| Target Compounds            |       |      |          |       |        |          |
| 3) Naphthalene              | 10.76 | 128  | 3023     | 0.097 | ng/ul# | 96       |
| 5) 2-Methylnaphthalene      | 12.37 | 142  | 577      | 0.025 | ng/ul  | 99       |
| 12) Phenanthrene            | 17.29 | 178  | 1869     | 0.035 | ng/ul# | 90       |
| 15) Fluoranthene            | 19.30 | 202  | 3916     | 0.058 | ng/ul  | 86       |
| 17) Pyrene                  | 19.67 | 202  | 4055     | 0.066 | ng/ul# | 88       |
| 18) Benzo(a)anthracene      | 21.39 | 228  | 2588     | 0.042 | ng/ul# | 88       |
| 19) Chrysene                | 21.45 | 228  | 2765     | 0.047 | ng/ul  | 92       |
| 21) Benzo(b)fluoranthene    | 23.14 | 252  | 4359     | 0.075 | ng/ul  | 87       |
| 23) Benzo(a)pyrene          | 23.80 | 252  | 3430m    | 0.062 | ng/ul  |          |
| 26) Benzo(g,h,i)perylene    | 27.36 | 276  | 2115m    | 0.041 | ng/ul  |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA\_E\DATA\BE072517\  
Data File : BE093589.D  
Acq On : 25 Jul 2017 22:21  
Operator : SJ/JU  
Sample : I4209-18  
Misc :  
ALS Vial : 15 Sample Multiplier: 1

Instrument :  
BNA\_E  
ClientSampleId :  
C0B21

Manual Integrations  
APPROVED

Sohil  
7/26/2017 5:56:07 PM

Quant Time: Jul 26 10:37:24 2017  
Quant Method : Z:\HPCHEM1\BNA\_E\METHODS\SOM-EPA-SIM-BE072417.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Tue Jul 25 06:10:24 2017  
Response via : Initial Calibration

