

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE072517\
 Data File : BE093596.D
 Acq On : 26 Jul 2017 2:49
 Operator : SJ/JU
 Sample : I4373-03
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 C0AA2

Manual Integrations
 APPROVED

Sohil
 7/26/2017 5:56:19 PM

Quant Time: Jul 26 06:08:37 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE072417.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 26 05:32:22 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.92	152	3089	0.40	ng/ul	0.00
2) Naphthalene-d8	10.71	136	16378	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.53	164	10496	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.26	188	24677m	0.40	ng/ul	0.00
16) Chrysene-d12	21.41	240	26139	0.40	ng/ul	0.00
20) Perylene-d12	23.91	264	22570m	0.40	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Methylnaphthalene-d10	12.29	152	6411	0.24	ng/ul	0.00
14) Fluoranthene-d10	19.28	212	21974	0.28	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE072517\
Data File : BE093596.D
Acq On : 26 Jul 2017 2:49
Operator : SJ/JU
Sample : I4373-03
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
C0AA2

Quant Time: Jul 26 06:08:37 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE072417.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 26 05:32:22 2017
Response via : Initial Calibration

Manual Integrations
APPROVED

Sohil
7/26/2017 5:56:19 PM

