

Data Path : Z:\HPCHEM1\BNA_E\Data\BE072617\
 Data File : BE093600.D
 Acq On : 26 Jul 2017 11:41
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 27 08:22:46 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE072117.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 21 16:13:20 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	82	-0.01
2	1,4-Dioxane	0.400	0.432	-8.0	90	0.00
3	n-Nitrosodimethylamine	0.400	0.387	3.3	86	0.01
4 S	2-Fluorophenol	0.400	0.417	-4.2	87	0.00
5 S	Phenol-d6	0.400	0.437	-9.2	94	-0.01
6	bis(2-Chloroethyl)ether	0.400	0.427	-6.7	90	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	84	0.00
8 S	Nitrobenzene-d5	0.400	0.426	-6.5	93	0.00
9	Naphthalene	0.400	0.431	-7.7	93	0.00
10	Hexachlorobutadiene	0.400	0.424	-6.0	90	-0.02
11 SURR	2-Methylnaphthalene-d10	0.400	0.437	-9.2	93	0.00
12	2-Methylnaphthalene	0.400	0.435	-8.7	94	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	84	0.00
14 S	2,4,6-Tribromophenol	0.400	0.269	32.8#	55	0.00
15 S	2-Fluorobiphenyl	0.400	0.440	-10.0	90	0.00
16	Acenaphthylene	0.400	0.425	-6.2	92	0.00
17	Acenaphthene	0.400	0.438	-9.5	92	0.00
18	Fluorene	0.400	0.414	-3.5	87	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	72	0.00
20	4-Bromophenyl-phenylether	0.400	0.404	-1.0	83	0.00
21	Hexachlorobenzene	0.400	0.411	-2.7	82	0.00
22	Pentachlorophenol	0.400	0.522	-30.5#	103	0.00
23	Phenanthrene	0.400	0.447	-11.7	85	0.00
24	Anthracene	0.400	0.433	-8.2	82	0.00
25 SURR	Fluoranthene-d10	0.400	0.339	15.3	64	0.00
26	Fluoranthene	0.400	0.341	14.8	65	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	57	-0.01
28	Pyrene	0.400	0.441	-10.2	64	0.00
29 S	Terphenyl-d14	0.400	0.429	-7.2	61	0.00
30	Benzo(a)anthracene	0.400	0.417	-4.2	61	0.00
31	Chrysene	0.400	0.443	-10.7	64	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.327	18.3	48	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.479	-19.7	70	0.00
34 I	Perylene-d12	0.400	0.400	0.0	64	0.00
35	Benzo(b)fluoranthene	0.400	0.409	-2.2	66	0.00
36	Benzo(k)fluoranthene	0.400	0.426	-6.5	70	0.00
37 C	Benzo(a)pyrene	0.400	0.419	-4.7	69	0.00
38	Dibenzo(a,h)anthracene	0.400	0.427	-6.7	70	0.00
39	Benzo(g,h,i)perylene	0.400	0.412	-3.0	67	0.00

Data Path : Z:\HPCHEM1\BNA_E\Data\BE072617\
Data File : BE093600.D
Acq On : 26 Jul 2017 11:41
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 27 08:22:46 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE072117.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 21 16:13:20 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				