

Data Path : Z:\HPCHEM1\BNA_E\Data\BE072617\
 Data File : BE093610.D
 Acq On : 26 Jul 2017 18:28
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 27 08:35:37 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE072117.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 21 16:13:20 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	83	0.00
2	1,4-Dioxane	0.400	0.413	-3.2	86	0.00
3	n-Nitrosodimethylamine	0.400	0.373	6.8	83	0.00
4 S	2-Fluorophenol	0.400	0.405	-1.3	85	0.00
5 S	Phenol-d6	0.400	0.439	-9.7	95	0.00
6	bis(2-Chloroethyl)ether	0.400	0.428	-7.0	91	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	83	0.00
8 S	Nitrobenzene-d5	0.400	0.436	-9.0	95	0.00
9	Naphthalene	0.400	0.437	-9.2	94	0.00
10	Hexachlorobutadiene	0.400	0.418	-4.5	89	-0.02
11 SURR	2-Methylnaphthalene-d10	0.400	0.456	-14.0	97	0.00
12	2-Methylnaphthalene	0.400	0.449	-12.2	96	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	91	0.00
14 S	2,4,6-Tribromophenol	0.400	0.282	29.5#	62	0.00
15 S	2-Fluorobiphenyl	0.400	0.429	-7.2	95	-0.02
16	Acenaphthylene	0.400	0.417	-4.2	98	0.00
17	Acenaphthene	0.400	0.425	-6.2	96	0.00
18	Fluorene	0.400	0.404	-1.0	91	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	85	0.00
20	4-Bromophenyl-phenylether	0.400	0.345	13.8	83	0.00
21	Hexachlorobenzene	0.400	0.369	7.8	86	0.00
22	Pentachlorophenol	0.400	0.478	-19.5	104	0.00
23	Phenanthrene	0.400	0.413	-3.2	91	0.00
24	Anthracene	0.400	0.404	-1.0	89	0.00
25 SURR	Fluoranthene-d10	0.400	0.311	22.3	68	0.00
26	Fluoranthene	0.400	0.313	21.8	69	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	61	-0.01
28	Pyrene	0.400	0.447	-11.7	69	0.00
29 S	Terphenyl-d14	0.400	0.438	-9.5	67	0.00
30	Benzo(a)anthracene	0.400	0.413	-3.2	64	0.00
31	Chrysene	0.400	0.440	-10.0	67	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.295	26.3#	46	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.471	-17.7	73	0.00
34 I	Perylene-d12	0.400	0.400	0.0	65	0.00
35	Benzo(b)fluoranthene	0.400	0.432	-8.0	70	0.00
36	Benzo(k)fluoranthene	0.400	0.444	-11.0	73	0.00
37 C	Benzo(a)pyrene	0.400	0.438	-9.5	73	0.00
38	Dibenzo(a,h)anthracene	0.400	0.444	-11.0	73	0.00
39	Benzo(g,h,i)perylene	0.400	0.427	-6.7	70	0.00

Data Path : Z:\HPCHEM1\BNA_E\Data\BE072617\
Data File : BE093610.D
Acq On : 26 Jul 2017 18:28
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 27 08:35:37 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE072117.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 21 16:13:20 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				